



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94882

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2322
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	3.99	04/25/2018	13:25
2	CAL2	1	Water			21.3	7.01	04/25/2018	13:26
3	CAL3	1	Water			21.3	10.02	04/25/2018	13:27
4	ICV	1	Water			21.2	7.00	04/25/2018	13:30
5	CCV1	1	Water			21.3	2.01	04/25/2018	13:31
6	J2557-01	1	Water			19.5	4.43	04/25/2018	13:33
7	J2557-01DUP	1	Water			19.6	4.44	04/25/2018	13:34
8	J2557-02	1	Water			19.8	4.57	04/25/2018	13:36
9	CCV2	1	Water			21.3	12.02	04/25/2018	13:37

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2257

WorkList ID : 111253

Date : 4/25/2018 1:18:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/23/2018	Water	J2557-01-A	pH	Cool 4 deg C	STER16	G41	MAIN-CELL	04/23/2018	SM 4500-PH B
04/23/2018	Water	J2557-02 -A	pH	Cool 4 deg C	STER16	G41	EX-CELL	04/23/2018	SM 4500-PH B

Date/Time 04/25/18 1:20 PM
 Received by: JD
 Relinquished by: C

Date/Time 04/25/18 2:10 PM
 Received by: C
 Relinquished by: JD