


7471B

LB 94888
INSTRUMENT ID: CVL

Linear

 μ Abs.:

45690



A= 0.0000e+000

B= 2.1674e-004

C= 1.2096e-002

Rho= 0.9997658

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	49	0.000	49				
0.2	0.200	0.218	0.018	949	0.0 %	949				
2.5	2.500	2.416	-0.084	11093	0.0 %	11093				
5.0	5.000	4.977	-0.023	22909	0.0 %	22909				
7.5	7.500	7.651	0.151	35243	0.0 %	35243				
10.0	10.000	9.915	-0.085	45690	0.0 %	45690				

LB94888 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Apr 2018 15:15:00

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	49	-	0.00007471B	PPB	25 Apr 2018 15:18:05	S	Std
0.2 - 1	50.2	949	-	0.20007471B	PPB	25 Apr 2018 15:20:23	S	Std
2.5 - 1	52.5	11093	-	2.50007471B	PPB	25 Apr 2018 15:22:40	S	Std
5 - 1	55	22909	-	7471B	PPB	25 Apr 2018 15:24:58	S	Std
7.5 - 1	57.5	35243	-	7.50007471B	PPB	25 Apr 2018 15:27:15	S	Std
10 - 1	510	45690	-	7471B	PPB	25 Apr 2018 15:32:13	S	Std
ICV14 - 1	ICV14	17345	3.7715	-7471B	PPB	25 Apr 2018 15:34:57	U	SMPL
ICB14 - 1	ICB14	-84	-0.0061	-7471B	PPB	25 Apr 2018 15:37:12	U	SMPL
CCV30 - 1	CCV30	23050	5.0080	-7471B	PPB	25 Apr 2018 15:39:28	U	SMPL
CCB30 - 1	CCB30	-130	-0.0161	-7471B	PPB	25 Apr 2018 15:41:44	U	SMPL
CRA - 1	CRA	945	0.2169	-7471B	PPB	25 Apr 2018 15:44:01	U	SMPL
HighStd - 1	HighStd	46838	10.1638	-7471B	PPB	25 Apr 2018 15:46:16	U	SMPL
ChkStd - 1	ChkStd	29954	6.5044	-7471B	PPB	25 Apr 2018 15:48:33	U	SMPL
PB108577BL - 1	PBS	-203	-0.0319	-7471B	PPB	25 Apr 2018 15:50:51	U	SMPL
PB108577BS - 1	LCSS	15024	3.2684	-7471B	PPB	25 Apr 2018 15:53:09	U	SMPL
J2155-07 - 1	NB-08-042418-A	1932	0.4308	-7471B	PPB	25 Apr 2018 15:55:26	U	SMPL
J2155-09 - 1	NB-08-042418-B	2725	0.6027	-7471B	PPB	25 Apr 2018 15:57:44	U	SMPL
J2155-11 - 1	NB-08-042418-C	1700	0.3806	-7471B	PPB	25 Apr 2018 16:00:01	U	SMPL
J2156-03 - 1	HD-02-042418	727	0.1697	-7471B	PPB	25 Apr 2018 16:02:19	U	SMPL
J2158-03 - 1	OK-01-042418	1815	0.4055	-7471B	PPB	25 Apr 2018 16:04:34	U	SMPL
CCV31 - 1	CCV31	22761	4.9453	-7471B	PPB	25 Apr 2018 16:06:50	U	SMPL
CCB31 - 1	CCB31	-96	-0.0087	-7471B	PPB	25 Apr 2018 16:09:06	U	SMPL
J2525-01 - 1	IM-TS-SJAP-01	3493	0.7692	-7471B	PPB	25 Apr 2018 16:11:21	U	SMPL
J2525-02 - 1	IM-TS-SJAP-02	4281	0.9400	-7471B	PPB	25 Apr 2018 16:13:37	U	SMPL
J2525-03 - 1	IM-TS-SJAP-03	2494	0.5526	-7471B	PPB	25 Apr 2018 16:15:54	U	SMPL
J2525-04 - 1	IM-TS-SJAP-04	2629	0.5819	-7471B	PPB	25 Apr 2018 16:18:10	U	SMPL
J2525-05 - 1	IM-TS-SJAP-05	4317	0.9478	-7471B	PPB	25 Apr 2018 16:20:27	U	SMPL
J2525-06 - 1	IM-TS-SJAP-06	3548	0.7811	-7471B	PPB	25 Apr 2018 16:22:44	U	SMPL
J2525-07 - 1	IM-TS-SJAP-07	3733	0.8212	-7471B	PPB	25 Apr 2018 16:25:00	U	SMPL
J2525-08 - 1	IM-TS-SJAP-08	4349	0.9547	-7471B	PPB	25 Apr 2018 16:27:18	U	SMPL
J2538-01 - 1	PAINT-CHIPS-DRUM-2-1	3402	0.7494	-7471B	PPB	25 Apr 2018 16:29:36	U	SMPL
J2539-01 - 1	FK-EF-01-005-ABCD	1550	0.3480	-7471B	PPB	25 Apr 2018 16:31:52	U	SMPL
CCV32 - 1	CCV32	23106	5.0201	-7471B	PPB	25 Apr 2018 16:34:08	U	SMPL
CCB32 - 1	CCB32	-101	-0.0098	-7471B	PPB	25 Apr 2018 16:36:24	U	SMPL
J2539-01DUP - 1	FK-EF-01-005-ABCD DUP	1595	0.3578	-7471B	PPB	25 Apr 2018 16:38:41	U	SMPL
J2539-01MS - 1	FK-EF-01-005-ABCDMS	19123	4.1568	-7471B	PPB	25 Apr 2018 16:40:58	U	SMPL
J2539-01MSD - 1	FK-EF-01-005-ABCDMSD	18972	4.1241	-7471B	PPB	25 Apr 2018 16:43:14	U	SMPL
J2541-01 - 1	TP-4-11	1018	0.2327	-7471B	PPB	25 Apr 2018 16:45:31	U	SMPL
J2539-01LX5 - 1		250	0.0663	-7471B	PPB	25 Apr 2018 16:47:48	U	SMPL
J2539-01A - 1		20117	4.3723	-7471B	PPB	25 Apr 2018 16:50:06	U	SMPL
CCV33 - 1	CCV33	22575	4.9050	-7471B	PPB	25 Apr 2018 16:52:23	U	SMPL
CCB33 - 1	CCB33	-152	-0.0208	-7471B	PPB	25 Apr 2018 16:54:41	U	SMPL