

===== LB94892 =====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS4/25/2018 15:46
-----Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.753	0.0	0.096	
ICB1	-0.129	0.0	0.004	
CCV1	243.520	0.0	0.237	
CCB1	-0.164	0.0	0.004	
PB108606BL	-0.334	0.0	0.004	
PB108606BS	96.930	0.0	0.097	
J2155-07	0.018	0.0	0.004	
J2155-07DUP	-0.202	0.0	0.004	
J2155-07MS	33.096	0.0	0.036	
J2155-07MSD	32.881	0.0	0.036	
J2155-09	1.463	0.0	0.006	
J2155-11	-0.868	0.0	0.003	
J2156-03	2.098	0.0	0.006	
J2158-03	1.019	0.0	0.005	
CCV2	250.763	0.0	0.244	
CCB2	-0.929	0.0	0.003	
J2539-01	-0.077	0.0	0.004	
J2541-01	-0.601	0.0	0.004	
CCV3	255.735	0.0	0.249	
CCB3	-0.632	0.0	0.004	

N	20
Mean	50.467
SD	91.0167
CV%	180.35

Aquakem v. 7.2AQ1

Results from time period:

Wed Apr 25 15:14:10 2018

Wed Apr 25 15:46:41 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.9136	µg/l	4/25/2018 8:33:03	
5.0PPBCN	A	Total CN	P	4.1459	µg/l	4/25/2018 8:33:04	
10PPBCN	A	Total CN	P	10.4022	µg/l	4/25/2018 8:33:05	
50PPBCN	A	Total CN	P	51.5898	µg/l	4/25/2018 8:33:06	
100PPBCN	A	Total CN	P	99.2618	µg/l	4/25/2018 8:33:07	
250PPBCN	A	Total CN	P	251.0494	µg/l	4/25/2018 8:33:08	
500PPBCN	A	Total CN	P	499.4645	µg/l	4/25/2018 8:33:09	
ICV1	S	Total CN	P	95.7533	µg/l	4/25/2018 15:14:10	
ICB1	S	Total CN	P	-0.1286	µg/l	4/25/2018 15:14:11	
CCV1	S	Total CN	P	243.5199	µg/l	4/25/2018 15:14:12	
CCB1	S	Total CN	P	-0.1644	µg/l	4/25/2018 15:14:13	
PB108606BL	S	Total CN	P	-0.3342	µg/l	4/25/2018 15:14:14	
PB108606BS	S	Total CN	P	96.9304	µg/l	4/25/2018 15:14:15	
J2155-07	S	Total CN	P	0.0176	µg/l	4/25/2018 15:14:16	
J2155-07DUP	S	Total CN	P	-0.2022	µg/l	4/25/2018 15:14:17	
J2155-07MS	S	Total CN	P	33.096	µg/l	4/25/2018 15:14:18	
J2155-07MSD	S	Total CN	P	32.8808	µg/l	4/25/2018 15:14:19	
J2155-09	S	Total CN	P	1.4635	µg/l	4/25/2018 15:14:20	
J2155-11	S	Total CN	P	-0.8675	µg/l	4/25/2018 15:20:49	
J2156-03	S	Total CN	P	2.0978	µg/l	4/25/2018 15:20:50	
J2158-03	S	Total CN	P	1.0191	µg/l	4/25/2018 15:20:51	
CCV2	S	Total CN	P	250.7632	µg/l	4/25/2018 15:46:36	
CCB2	S	Total CN	P	-0.9288	µg/l	4/25/2018 15:46:37	
J2539-01	S	Total CN	P	-0.0767	µg/l	4/25/2018 15:46:38	
J2541-01	S	Total CN	P	-0.601	µg/l	4/25/2018 15:46:39	
CCV3	S	Total CN	P	255.7347	µg/l	4/25/2018 15:46:40	
CCB3	S	Total CN	P	-0.6316	µg/l	4/25/2018 15:46:41	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

4/25/2018 14:57


Test Total CN

Accepted 4/25/2018 9:25

Factor 1046
 Bias 0.004

Coeff. of det. 0.999970

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	0.0PPBCN	0.003	-0.9136	0.0000	
2	5.0PPBCN	0.008	4.1459	5.0000	
3	10PPBCN	0.014	10.4022	10.0000	
4	50PPBCN	0.054	51.5898	50.0000	
5	100PPBCN	0.099	99.2618	100.0000	
6	250PPBCN	0.244	251.0494	250.0000	
7	500PPBCN	0.482	499.4645	500.0000	