

245.1


LB94895

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34220



A= 0.0000e+000

B= 2.9123e-004

C= -1.7364e-003

Rho= 0.9999649

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	-96	0.000	-96				
0.05	0.050	0.057	0.007	201	0.0 %	201				
0.2	0.200	0.187	-0.013	647	0.0 %	647				
2.5	2.500	2.504	0.004	8605	0.0 %	8605				
5	5.000	5.066	0.066	17401	0.0 %	17401				
7.5	7.500	7.502	0.002	25766	0.0 %	25766				
10	10.000	9.964	-0.036	34220	0.0 %	34220				

LB94895 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Apr 2018 17:13:35

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-96	-	0.0000	245.1	PPB	25 Apr 2018 17:30:30	S	Std
0.05 - 1	50.05	201	-	0.0500	245.1	PPB	25 Apr 2018 17:32:39	S	Std
0.2 - 1	50.2	647	-	0.2000	245.1	PPB	25 Apr 2018 17:34:48	S	Std
2.5 - 1	52.5	8605	-	2.5000	245.1	PPB	25 Apr 2018 17:36:56	S	Std
5 - 1	55	17401	-	5.0000	245.1	PPB	25 Apr 2018 17:39:04	S	Std
7.5 - 1	57.5	25766	-	7.5000	245.1	PPB	25 Apr 2018 17:41:11	S	Std
10 - 1	510	34220	-	10.0000	245.1	PPB	25 Apr 2018 17:43:18	S	Std
ICV15 - 1	ICV15	13368	3.8914	-245.1		PPB	25 Apr 2018 17:45:59	U	SMPL
ICB15 - 1	ICB15	-204	-0.0611	-245.1		PPB	25 Apr 2018 17:48:05	U	SMPL
CCV34 - 1	CCV34	17415	5.0700	-245.1		PPB	25 Apr 2018 17:50:12	U	SMPL
CCB34 - 1	CCB34	-201	-0.0603	-245.1		PPB	25 Apr 2018 17:52:19	U	SMPL
CRA - 1	CRA	198	0.0559	-245.1		PPB	25 Apr 2018 17:54:27	U	SMPL
HighStd - 1	HighStd	33793	9.8397	-245.1		PPB	25 Apr 2018 17:56:33	U	SMPL
ChkStd - 1	ChkStd	22472	6.5427	-245.1		PPB	25 Apr 2018 17:58:40	U	SMPL
PB108591BL - 1	PBW	-242	-0.0722	-245.1		PPB	25 Apr 2018 18:00:49	U	SMPL
PB108591BS - 1	LCSW	13686	3.9840	-245.1		PPB	25 Apr 2018 18:02:58	U	SMPL
J2531-01 - 1	042318-DBRI-E-D-B	-189	-0.0568	-245.1		PPB	25 Apr 2018 18:05:06	U	SMPL
J2531-02 - 1	042318-DBRI-E-D-S	-83	-0.0259	-245.1		PPB	25 Apr 2018 18:07:14	U	SMPL
J2531-03 - 1	042318-DBRI-E-U-B	-57	-0.0183	-245.1		PPB	25 Apr 2018 18:09:22	U	SMPL
J2531-04 - 1	042318-DBRI-E-U-S	-80	-0.0250	-245.1		PPB	25 Apr 2018 18:11:31	U	SMPL
J2531-05 - 1	042318-SIDI-E-D-M	-55	-0.0178	-245.1		PPB	25 Apr 2018 18:13:37	U	SMPL
CCV35 - 1	CCV35	17452	5.0807	-245.1		PPB	25 Apr 2018 18:15:44	U	SMPL
CCB35 - 1	CCB35	-200	-0.0600	-245.1		PPB	25 Apr 2018 18:17:51	U	SMPL
J2531-06 - 1	042318-SIDI-E-U-M	-82	-0.0256	-245.1		PPB	25 Apr 2018 18:19:58	U	SMPL
J2531-06DUP - 1	042318-SIDI-E-U-MDUP	-44	-0.0146	-245.1		PPB	25 Apr 2018 18:22:05	U	SMPL
J2531-06MS - 1	042318-SIDI-E-U-MMS	15615	4.5458	-245.1		PPB	25 Apr 2018 18:24:12	U	SMPL
J2531-06MSD - 1	042318-SIDI-E-U-MMSD	15236	4.4354	-245.1		PPB	25 Apr 2018 18:26:19	U	SMPL
J2531-06LX5 - 1		-145	-0.0440	-245.1		PPB	25 Apr 2018 18:28:26	U	SMPL
J2531-06A - 1		16106	4.6888	-245.1		PPB	25 Apr 2018 18:30:34	U	SMPL
CCV36 - 1	CCV36	17404	5.0668	-245.1		PPB	25 Apr 2018 18:32:42	U	SMPL
CCB36 - 1	CCB36	-209	-0.0626	-245.1		PPB	25 Apr 2018 18:34:51	U	SMPL