



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94903

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2322
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
Buffer Solution, PH12 (500ml)	W2267
pH 2.00 Buffer	W2289

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.00	04/26/2018	08:30
2	CAL2	1	Water			21.6	7.00	04/26/2018	08:32
3	CAL3	1	Water			21.4	10.03	04/26/2018	08:34
4	ICV	1	Water			21.3	7.00	04/26/2018	08:36
5	CCV1	1	Water			21.6	2.00	04/26/2018	08:38
6	J2573-01	1	Solid	20.3	20	20.6	6.19	04/26/2018	08:40
7	J2573-01DUP	1	Solid	20.1	20	20.7	6.20	04/26/2018	08:42
8	J2573-02	1	Solid	20.5	20	20.5	4.89	04/26/2018	08:44
9	J2573-03	1	Solid	20.2	20	20.4	4.77	04/26/2018	08:46
10	J2576-01	1	Solid	20.4	20	20.7	6.02	04/26/2018	08:48
11	J2579-01	1	Solid	20.2	20	20.4	6.18	04/26/2018	08:50
12	J2582-01	1	Solid	20.3	20	20.5	6.05	04/26/2018	08:52
13	J2585-01	1	Solid	20.4	20	20.3	5.98	04/26/2018	08:54
14	CCV2	1	Water			21.4	12.00	04/26/2018	08:56

WORKLIST(Hardcopy Internal Chain)

LB 94903

WorkList Name : ph042618

WorkList ID : 111285

Date : 4/26/2018 8:20:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/25/2018	Solid	J2573-01	pH	Cool 4 deg C	TULL02	G51	EME-4-25	04/25/2018	9045C
04/25/2018	Solid	J2573-02	pH	Cool 4 deg C	TULL02	G51	EME-BW	04/25/2018	9045C
04/25/2018	Solid	J2573-03	pH	Cool 4 deg C	TULL02	G51	EME-JCS	04/25/2018	9045C
04/25/2018	Solid	J2576-01	pH	Cool 4 deg C	TULL02	G62	FK-PS-01.023	04/25/2018	9045C
04/25/2018	Solid	J2579-01	pH	Cool 4 deg C	TULL02	G61	FK-PS-01-023	04/25/2018	9045C
04/25/2018	Solid	J2582-01	pH	Cool 4 deg C	TULL02	G52	FK-PS-01-022	04/25/2018	9045C
04/25/2018	Solid	J2585-01	pH	Cool 4 deg C	TULL02	G51	FK-PS-01.022	04/25/2018	9045C

Date/Time 04/26/18 8:25 AM
 Received by: SP
 Relinquished by: CP

Date/Time 04/26/18 9:15 AM
 Received by: CP
 Relinquished by: SP