



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: jigneshDate: 04/26/2018TEMP1 IN: 103 04/26/2018 07:35 TEMP OUT: 103 04/26/2018 08:35TEMP2 IN: 103 04/26/2018 09:15 TEMP OUT: 104 04/26/2018 10:20TEMP3 IN: 104 04/26/2018 11:15 TEMP OUT: 104 04/26/2018 12:40TEMP4 IN: 104 04/26/2018 13:20 TEMP OUT: 104 04/26/2018 14:15Run Number: LB94905BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB94905BL	LB94905BL	1.4366	1.4366	100	1.4366	1.4366	1.4366	0.0000	0
2	LB94905BS	LB94905BS	1.3066	1.3066	100	1.3611	1.3611	1.3611	0.0545	545
3	J2568-01	001M-WILLETS-PT-BLVD (MAR)	1.4027	1.4027	200	1.4091	1.4091	1.4091	0.0064	32
4	J2568-01DUP	001M-WILLETS-PT-BLVD (MAR) DU	1.4177	1.4177	200	1.4239	1.4239	1.4239	0.0062	31
5	J2568-02	002M-35TH-AVE (MAR)	1.4049	1.4049	200	1.4119	1.4119	1.4119	0.0070	35

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB94905

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS J2568

WorkList ID : 111292

Date : 4/26/2018 7:30:10 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2018	Water	J2568-01	TSS	Cool 4 deg C	TULL01	G62	001M-WILLETTS-PT-BLVD(MA	04/24/2018	SM2540 D
04/26/2018	Water	J2568-02	TSS	Cool 4 deg C	TULL01	G62	002M-35TH-AVE(MAR)	04/24/2018	SM2540 D

Date/Time 6/12/18 7:32 AM
 Received by: J
 Relinquished by:

Date/Time 6/12/18 10:00 AM
 Received by: J
 Relinquished by: J