

LB 94906


7470A

INSTRUMENT ID: CY2

Linear

 μ Abs.:

35390



A= 0.0000e+000

B= 2.8232e-004

C= -1.5832e-002

Rho= 0.9999837

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	-9	0.000	-9				
0.2	0.200	0.183	-0.017	706	0.0 %	706				
2.5	2.500	2.527	0.027	9007	0.0 %	9007				
5.0	5.000	5.026	0.026	17858	0.0 %	17858				
7.5	7.500	7.507	0.007	26646	0.0 %	26646				
10.0	10.000	9.975	-0.025	35390	0.0 %	35390				

LB94906 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Apr 2018 10:26:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-9	-	0.00007470A	PPB		26 Apr 2018 10:30:25	S	Std
0.2 - 1	50-2	706	-	0.20007470A	PPB		26 Apr 2018 10:32:33	S	Std
2.5 - 1	52-5	9007	-	2.50007470A	PPB		26 Apr 2018 10:34:42	S	Std
5.0 - 1	55	17858	-	5.00007470A	PPB		26 Apr 2018 10:36:50	S	Std
7.5 - 1	57-5	26646	-	7.50007470A	PPB		26 Apr 2018 10:38:58	S	Std
10.0 - 1	510	35390	-	10.00007470A	PPB		26 Apr 2018 10:41:06	S	Std
ICV16 - 1	ICV16	13952	3.9230	-7470A	PPB		26 Apr 2018 10:43:39	U	SMPL
ICB16 - 1	ICB16	-207	-0.0743	-7470A	PPB		26 Apr 2018 10:45:46	U	SMPL
CCV37 - 1	CCV37	18261	5.1395	-7470A	PPB		26 Apr 2018 10:47:53	U	SMPL
CCB37 - 1	CCB37	-183	-0.0675	-7470A	PPB		26 Apr 2018 10:50:00	U	SMPL
CRA - 1	CRA	697	0.1809	-7470A	PPB		26 Apr 2018 10:52:08	U	SMPL
HighStd - 1	HighStd	35160	9.9104	-7470A	PPB		26 Apr 2018 10:54:14	U	SMPL
ChkStd - 1	ChkStd	24564	6.9190	-7470A	PPB		26 Apr 2018 10:56:21	U	SMPL
PB108609BL - 1	PBW	-238	-0.0830	-7470A	PPB		26 Apr 2018 10:58:30	U	SMPL
PB108609BS - 1	LCSW	14019	3.9420	-7470A	PPB		26 Apr 2018 11:00:39	U	SMPL
PB108609TB - 1	PB108568TB	-131	-0.0528	-7470A	PPB		26 Apr 2018 11:02:47	U	SMPL
J2155-08 - 1	NB-08-042418-A	0	-0.0158	-7470A	PPB		26 Apr 2018 11:04:55	U	SMPL
J2155-10 - 1	NB-08-042418-B	-3	-0.0167	-7470A	PPB		26 Apr 2018 11:07:03	U	SMPL
J2155-12 - 1	NB-08-042418-C	0	-0.0158	-7470A	PPB		26 Apr 2018 11:09:11	U	SMPL
J2156-04 - 1	HD-02-042418	-14	-0.0198	-7470A	PPB		26 Apr 2018 11:11:18	U	SMPL
CCV38 - 1	CCV38	18611	5.2384	-7470A	PPB		26 Apr 2018 11:13:24	U	SMPL
CCB38 - 1	CCB38	-146	-0.0571	-7470A	PPB		26 Apr 2018 11:15:31	U	SMPL
J2158-04 - 1	OK-01-042418	1	-0.0156	-7470A	PPB		26 Apr 2018 11:17:39	U	SMPL
J2158-04DUP - 1	OK-01-042418DUP	37	-0.0054	-7470A	PPB		26 Apr 2018 11:19:46	U	SMPL
J2158-04MS - 1	OK-01-042418MS	13059	3.6709	-7470A	PPB		26 Apr 2018 11:21:53	U	SMPL
J2158-04MSD - 1	OK-01-042418MSD	13316	3.7435	-7470A	PPB		26 Apr 2018 11:24:00	U	SMPL
J2538-02 - 1	PAINT-CHIPS-DRUM-2-1	60	0.0011	-7470A	PPB		26 Apr 2018 11:28:16	U	SMPL
J2541-02 - 1	TP-3-4-5-WC-1	-5	-0.0172	-7470A	PPB		26 Apr 2018 11:30:24	U	SMPL
PB108610BL - 1	PBW	17	-0.0110	-7470A	PPB		26 Apr 2018 11:32:32	U	SMPL
PB108610BS - 1	LCSW	13436	3.7774	-7470A	PPB		26 Apr 2018 11:34:41	U	SMPL
PB108610TB - 1	PB108573TB	-74	-0.0367	-7470A	PPB		26 Apr 2018 11:36:48	U	SMPL
CCV39 - 1	CCV39	18609	5.2378	-7470A	PPB		26 Apr 2018 11:38:55	U	SMPL
CCB39 - 1	CCB39	-140	-0.0554	-7470A	PPB		26 Apr 2018 11:41:02	U	SMPL
J2543-02 - 1	TOTE	-11	-0.0189	-7470A	PPB		26 Apr 2018 11:43:10	U	SMPL
J2543-02DUP - 1	TOTEDUP	0	-0.0158	-7470A	PPB		26 Apr 2018 11:45:17	U	SMPL
J2543-02MS - 1	TOTEMS	13491	3.7929	-7470A	PPB		26 Apr 2018 11:47:24	U	SMPL
J2543-02MSD - 1	TOTEMSD	13481	3.7901	-7470A	PPB		26 Apr 2018 11:49:32	U	SMPL
PB108611BL - 1	PBW	-133	-0.0534	-7470A	PPB		26 Apr 2018 11:51:40	U	SMPL
PB108611BS - 1	LCSW	13691	3.8494	-7470A	PPB		26 Apr 2018 11:53:48	U	SMPL
J2543-01 - 1	TOTE	426	0.1044	-7470A	PPB		26 Apr 2018 12:00:30	U	SMPL
J2543-01DUP - 1	TOTEDUP	375	0.0900	-7470A	PPB		26 Apr 2018 12:02:38	U	SMPL
J2543-01MS - 1	TOTEMS	15699	4.4163	-7470A	PPB		26 Apr 2018 12:04:46	U	SMPL
J2543-01MSD - 1	TOTEMSD	16519	4.6478	-7470A	PPB		26 Apr 2018 12:06:53	U	SMPL
CCV40 - 1	CCV40	17883	5.0328	-7470A	PPB		26 Apr 2018 12:09:01	U	SMPL
CCB40 - 1	CCB40	-130	-0.0525	-7470A	PPB		26 Apr 2018 12:11:08	U	SMPL
J2158-04LX5 - 1		2	-0.0153	-7470A	PPB		26 Apr 2018 12:13:16	U	SMPL
J2158-04A - 1		13810	3.8830	-7470A	PPB		26 Apr 2018 12:15:24	U	SMPL
J2543-02LX5 - 1		-124	-0.0508	-7470A	PPB		26 Apr 2018 12:17:31	U	SMPL
J2543-02A - 1		13802	3.8807	-7470A	PPB		26 Apr 2018 12:19:38	U	SMPL
J2543-01LX5 - 1		-43	-0.0280	-7470A	PPB		26 Apr 2018 12:21:45	U	SMPL
J2543-01A - 1		16625	4.6777	-7470A	PPB		26 Apr 2018 12:23:53	U	SMPL
CCV41 - 1	CCV41	18174	5.1150	-7470A	PPB		26 Apr 2018 12:26:01	U	SMPL
CCB41 - 1	CCB41	-152	-0.0587	-7470A	PPB		26 Apr 2018 12:28:11	U	SMPL