


7471B

LB94909
INSTRUMENT ID: CY1

Linear

 μ Abs.:

44975



A= 0.0000e+000

B= 2.2264e-004

C= -1.6910e-003

Rho= 0.9999980

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	32	0.000	32				
0.2	0.200	0.202	0.002	915	0.0 %	915				
2.5	2.500	2.497	-0.003	11223	0.0 %	11223				
5.0	5.000	4.995	-0.005	22445	0.0 %	22445				
7.5	7.500	7.488	-0.012	33642	0.0 %	33642				
10.0	10.000	10.012	0.012	44975	0.0 %	44975				

LB94909 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Apr 2018 13:01:45

Sample ID	Extended ID	μ Abs.	Conc.	Stdnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	7471B	PPB	26 Apr 2018 13:04:05	S	Std
0.2 - 1	50.2	915	-	0.2000	7471B	PPB	26 Apr 2018 13:06:23	S	Std
2.5 - 1	52.5	11223	-	2.5000	7471B	PPB	26 Apr 2018 13:08:40	S	Std
5.0 - 1	55	22445	-	5.0000	7471B	PPB	26 Apr 2018 13:10:57	S	Std
7.5 - 1	57.5	33642	-	7.5000	7471B	PPB	26 Apr 2018 13:15:51	S	Std
10.0 - 1	510	44975	-	10.0000	7471B	PPB	26 Apr 2018 13:18:08	S	Std
ICV17 - 1	ICV17	16586	3.6910	-	7471B	PPB	26 Apr 2018 13:21:07	U	SMPL
ICB17 - 1	ICB17	-141	-0.0331	-	7471B	PPB	26 Apr 2018 13:23:22	U	SMPL
CCV42 - 1	CCV42	21392	4.7611	-	7471B	PPB	26 Apr 2018 13:25:39	U	SMPL
CCB42 - 1	CCB42	-159	-0.0371	-	7471B	PPB	26 Apr 2018 13:27:55	U	SMPL
CRA - 1	CRA	902	0.1991	-	7471B	PPB	26 Apr 2018 13:30:12	U	SMPL
HighStd - 1	HighStd	45023	10.0223	-	7471B	PPB	26 Apr 2018 13:32:28	U	SMPL
ChkStd - 1	ChkStd	30418	6.7706	-	7471B	PPB	26 Apr 2018 13:34:44	U	SMPL
PB108619BL - 1	PBS	-246	-0.0565	-	7471B	PPB	26 Apr 2018 13:37:03	U	SMPL
PB108619BS - 1	LCSS	16532	3.6790	-	7471B	PPB	26 Apr 2018 13:48:01	U	SMPL
J2154-01 - 1	JC-01-042518	3249	0.7217	-	7471B	PPB	26 Apr 2018 13:50:18	U	SMPL
J2156-05 - 1	HD-05-042518-A	5014	1.1146	-	7471B	PPB	26 Apr 2018 13:52:35	U	SMPL
J2156-07 - 1	HD-05-042518-B	2912	0.6466	-	7471B	PPB	26 Apr 2018 13:54:53	U	SMPL
J2156-09 - 1	HD-05-042518-C	4852	1.0786	-	7471B	PPB	26 Apr 2018 13:57:11	U	SMPL
J2156-09DUP - 1	HD-05-042518-CDUP	4768	1.0599	-	7471B	PPB	26 Apr 2018 13:59:27	U	SMPL
CCV43 - 1	CCV43	22304	4.9641	-	7471B	PPB	26 Apr 2018 14:01:43	U	SMPL
CCB43 - 1	CCB43	-118	-0.0280	-	7471B	PPB	26 Apr 2018 14:03:59	U	SMPL
J2156-09MS - 1	HD-05-042518-CMS	21750	4.8408	-	7471B	PPB	26 Apr 2018 14:06:16	U	SMPL
J2156-09MSD - 1	HD-05-042518-CMSD	21428	4.7691	-	7471B	PPB	26 Apr 2018 14:08:31	U	SMPL
J2161-07 - 1	OR-01-042518-A	4598	1.0220	-	7471B	PPB	26 Apr 2018 14:13:05	U	SMPL
J2161-09 - 1	OR-01-042518-B	1859	0.4122	-	7471B	PPB	26 Apr 2018 14:15:22	U	SMPL
J2549-01 - 1	SP-A	2239	0.4968	-	7471B	PPB	26 Apr 2018 14:17:39	U	SMPL
J2549-05 - 1	SP-B	2144	0.4757	-	7471B	PPB	26 Apr 2018 14:19:57	U	SMPL
J2549-09 - 1	SP-C	3593	0.7983	-	7471B	PPB	26 Apr 2018 14:22:14	U	SMPL
J2549-13 - 1	SP-D	3641	0.8089	-	7471B	PPB	26 Apr 2018 14:24:32	U	SMPL
J2564-01 - 1	LS-UB-27176	14734	3.2787	-	7471B	PPB	26 Apr 2018 14:26:48	U	SMPL
CCV44 - 1	CCV44	22208	4.9427	-	7471B	PPB	26 Apr 2018 14:29:05	U	SMPL
CCB44 - 1	CCB44	-115	-0.0273	-	7471B	PPB	26 Apr 2018 14:31:21	U	SMPL
J2565-01 - 1	TP-4_0.0-0.5	14677	3.2660	-	7471B	PPB	26 Apr 2018 14:33:38	U	SMPL
J2565-02 - 1	TP-4_2.0-2.5	886	0.1956	-	7471B	PPB	26 Apr 2018 14:35:55	U	SMPL
J2565-04 - 1	TP-3_0.0-0.5	13365	2.9739	-	7471B	PPB	26 Apr 2018 14:38:11	U	SMPL
J2565-05 - 1	TP-3_2.0-2.5	2805	0.6228	-	7471B	PPB	26 Apr 2018 14:40:28	U	SMPL
J2565-07 - 1	TP-2_0.0-0.5	14190	3.1576	-	7471B	PPB	26 Apr 2018 14:42:44	U	SMPL
J2565-08 - 1	TP-2_2.0-2.5	736	0.1622	-	7471B	PPB	26 Apr 2018 14:45:01	U	SMPL
J2565-10 - 1	TP-1_0.0-0.5	19892	4.4271	-	7471B	PPB	26 Apr 2018 14:47:18	U	SMPL
J2565-11 - 1	TP-1_2.0-2.5	7050	1.5679	-	7471B	PPB	26 Apr 2018 14:49:35	U	SMPL
PB108620BL - 1	PBS	-53	-0.0135	-	7471B	PPB	26 Apr 2018 14:51:53	U	SMPL
PB108620BS - 1	LCSS	15749	3.5047	-	7471B	PPB	26 Apr 2018 14:54:10	U	SMPL
CCV45 - 1	CCV45	21990	4.8942	-	7471B	PPB	26 Apr 2018 14:56:26	U	SMPL
CCB45 - 1	CCB45	-129	-0.0304	-	7471B	PPB	26 Apr 2018 14:58:42	U	SMPL
J2566-01 - 1	FRONT	5116	1.1373	-	7471B	PPB	26 Apr 2018 15:00:59	U	SMPL
J2566-01DUP - 1	FRONTDUP	4491	0.9982	-	7471B	PPB	26 Apr 2018 15:03:15	U	SMPL
J2566-01MS - 1	FRONTMS	21029	4.6802	-	7471B	PPB	26 Apr 2018 15:05:32	U	SMPL
J2566-01MSD - 1	FRONTMSD	20881	4.6473	-	7471B	PPB	26 Apr 2018 15:07:48	U	SMPL
J2566-02 - 1	BACK	12046	2.6802	-	7471B	PPB	26 Apr 2018 15:10:05	U	SMPL
J2573-01 - 1	EME-4-25	1697	0.3761	-	7471B	PPB	26 Apr 2018 15:12:22	U	SMPL
J2573-02 - 1	EME-BW	3818	0.8484	-	7471B	PPB	26 Apr 2018 15:14:39	U	SMPL
J2573-03 - 1	EME-JCS	260	0.0562	-	7471B	PPB	26 Apr 2018 15:16:57	U	SMPL
J2576-01 - 1	FK-PS-01.023	1381	0.3058	-	7471B	PPB	26 Apr 2018 15:19:15	U	SMPL
J2579-01 - 1	FK-PS-01-023	2168	0.4810	-	7471B	PPB	26 Apr 2018 15:21:31	U	SMPL
CCV46 - 1	CCV46	22058	4.9093	-	7471B	PPB	26 Apr 2018 15:23:48	U	SMPL
CCB46 - 1	CCB46	-99	-0.0237	-	7471B	PPB	26 Apr 2018 15:26:04	U	SMPL
J2582-01 - 1	FK-PS-01-022	1458	0.3229	-	7471B	PPB	26 Apr 2018 15:28:21	U	SMPL
J2585-01 - 1	FK-PS-01.022	5119	1.1380	-	7471B	PPB	26 Apr 2018 15:30:37	U	SMPL
J2590-01 - 1	EPE-F-4A(0-5)	10362	2.3053	-	7471B	PPB	26 Apr 2018 15:32:53	U	SMPL
J2590-02 - 1	EPE-F-4A(5-10)	965	0.2132	-	7471B	PPB	26 Apr 2018 15:35:10	U	SMPL
J2590-03 - 1	EPE-MX-4(0-5)	4837	1.0752	-	7471B	PPB	26 Apr 2018 15:37:27	U	SMPL
J2590-04 - 1	EPE-MX-4(5-10)	3053	0.6780	-	7471B	PPB	26 Apr 2018 15:39:43	U	SMPL
J2156-09LX5 - 1		894	0.1974	-	7471B	PPB	26 Apr 2018 15:42:01	U	SMPL
J2156-09A - 1		23683	5.2711	-	7471B	PPB	26 Apr 2018 15:44:18	U	SMPL
J2566-01LX5 - 1		831	0.1833	-	7471B	PPB	26 Apr 2018 15:46:36	U	SMPL
J2566-01A - 1		21868	4.8670	-	7471B	PPB	26 Apr 2018 15:48:53	U	SMPL
CCV47 - 1	CCV47	22196	4.9401	-	7471B	PPB	26 Apr 2018 15:51:10	U	SMPL
CCB47 - 1	CCB47	-126	-0.0297	-	7471B	PPB	26 Apr 2018 15:53:27	U	SMPL