



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB94915
Analysis Date: 04/26/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: Sweta
REVIEWED BY: apatel
Extraction Date: 04/26/2018
Extraction IN Time: 12:30
Extraction OUT Time: 13:10

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB94915BL	LB94915BL	WATER	2	1000	100	2.8921	2.8921	3.02	2.8921	2.8921	0.0000	0
2	LB94915BS	LB94915BS	WATER	2	1000	100	3.0135	3.0135	3.00	3.0315	3.0315	0.0180	18
3	J2563-01	STORM-DRAIN-1	WATER	2	1000	100	3.1023	3.1023	3.01	3.1026	3.1026	0.0003	0.3
4	J2563-01DUP	STORM-DRAIN-1DUP	WATER	2	1000	100	3.1024	3.1024	3.02	3.1028	3.1028	0.0004	0.4
5	J2563-02	STORM-DRAIN-2	WATER	2	1000	100	2.9600	2.9600	3.04	2.9603	2.9603	0.0003	0.3
6	J2593-01	EFF-180425-07	WATER	2	1000	100	2.9892	2.9892	3.00	2.9893	2.9893	0.0001	0.1
7	J2593-01MS	EFF-180425-07	WATER	2	1000	100	2.9893	2.9893	3.04	3.0096	3.0096	0.0203	20.3
8	J2593-01MSD	EFF-180425-07	WATER	2	1000	100	2.9898	2.9898	3.04	3.0102	3.0102	0.0204	20.4



QC Batch# LB94915 Test: TPH Analysis Date: 04/26/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2341
pH Paper 0-14	W2248
Sodium Sulfate	EP1839
1:1 HCL	WP59249
Silica Gel	W2199
Sand	N/A

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58005
LCSWD	N/A	N/A
MS/MSD	10.00 ML	WP58006

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>14:36</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>13:32</u>		
Bal Check Time:	<u>12:40</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>15:10</u>
		Out Time:	<u>14:35</u>		

After Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>16:11</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>15:36</u>		
Bal Check Time:	<u>17:05</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>16:38</u>
		Out Time:	<u>16:10</u>		

WORKLIST(Hardcopy Internal Chain)

LB 94915

Worklist Name : TPH042618

Worklist ID : 111295

Date : 4/26/2018 12:10:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2018	Water	J2563-01	E ₁ F ₁ TPH	Conc H2SO4 tc	KWAR01	G52	STORM-DRAIN-1	04/25/2018	1664A
05/02/2018	Water	J2563-02	E ₁ F ₁ TPH	Conc H2SO4 tc	KWAR01	G52	STORM-DRAIN-2	04/25/2018	1664A
04/29/2018	Water	J2593-01	E ₁ C ₁ TPH	Conc H2SO4 tc	VIAS01	H62	EFF-180425-07	04/25/2018	1664A

H₁I

Date/Time

04/26/18 12:15 PM

Received by:

SP

Relinquished by:

SP

Date/Time

04/26/18 2:00 PM

Received by:

SP

Relinquished by:

SP