



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94924

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2322
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	04/26/2018	16:10
2	CAL2	1	Water			21.2	7.00	04/26/2018	16:12
3	CAL3	1	Water			21.5	10.03	04/26/2018	16:14
4	ICV	1	Water			21.5	7.00	04/26/2018	16:16
5	CCV1	1	Water			21.4	2.00	04/26/2018	16:18
6	J2597-03	1	Solid	20.3	20	21.5	8.56	04/26/2018	16:20
7	J2597-03DUP	1	Solid	20.1	20	21.4	8.57	04/26/2018	16:22
8	J2608-02	1	Solid	20.5	20	21.2	10.34	04/26/2018	16:24
9	J2610-01	1	Solid	20.2	20	21.4	7.24	04/26/2018	16:26
10	CCV2	1	Water			21.5	12.00	04/26/2018	16:28

WORKLIST(Hardcopy Internal Chain)

LB 94924

Worklist Name : Corrosivity042618

Worklist ID : 111318

Date : 4/26/2018 3:51:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2018	Solid	J2597-03	Corrosivity	Cool 4 deg C	PSEG01	H21	KS-RO-133	04/26/2018	9045C
04/26/2018	Solid	J2608-02	Corrosivity	Cool 4 deg C	PSEG01	H22	CO-5-1	04/26/2018	9045C
04/26/2018	Solid	J2610-01	Corrosivity	Cool 4 deg C	BEMS01	H11	SB-01-WC	04/26/2018	9045C

Date/Time

04/26/18 4:05 PM

Received by:

SP

Relinquished by:

CP

Date/Time

04/26/18 4:40 PM

Received by:

CP

Relinquished by:

SP