



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94925

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2322
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	04/26/2018	16:10
2	CAL2	1	Water			21.2	7.00	04/26/2018	16:12
3	CAL3	1	Water			21.5	10.03	04/26/2018	16:14
4	ICV	1	Water			21.5	7.00	04/26/2018	16:16
5	CCV1	1	Water			21.4	2.00	04/26/2018	16:18
6	J2564-06	1	Water			20.4	7.32	04/26/2018	16:33
7	J2564-06DUP	1	Water			20.5	7.33	04/26/2018	16:35
8	CCV2	1	Water			21.3	12.00	04/26/2018	16:37

WORKLIST(Hardcopy Internal Chain)

LB 94925

WorkList Name : phj2564

WorkList ID : 111319

Date : 4/26/2018 3:53:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2018	Water	J2564-06	pH	Cool 4 deg C	PSEG01		LS-UB-102	04/26/2018	9040C

Date/Time

04/26/18 4:05 PM

Received by:

SP

Relinquished by:

SP

Date/Time

04/26/18 4:40 PM

Received by:

SP

Relinquished by:

SP