

245.1


LB94928

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35371



A= 0.0000e+000

B= 2.7752e-004

C= -4.9622e-003

Rho= 0.9993039

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-16	0.000	-16				
0.05	0.050	0.046	-0.004	184	0.0 %	184				
0.2	0.200	0.192	-0.008	709	0.0 %	709				
2.5	2.500	2.515	0.015	9081	0.0 %	9081				
5.0	5.000	4.898	-0.102	17668	0.0 %	17668				
7.5	7.500	7.797	0.297	28111	0.0 %	28111				
10.0	10.000	9.811	-0.189	35371	0.0 %	35371				

LB94928 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 26 Apr 2018 16:33:56

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	S0	-16	-	0.0000245.1	PPB	26 Apr 2018 16:38:23	S	Std
0.05 - 1	S0.05	184	-	0.0500245.1	PPB	26 Apr 2018 16:40:31	S	Std
0.2 - 1	S0.2	709	-	0.2000245.1	PPB	26 Apr 2018 16:42:40	S	Std
2.5 - 1	S2.5	9081	-	2.5000245.1	PPB	26 Apr 2018 16:44:48	S	Std
5 - 1	S5	17668	-	245.1	PPB	26 Apr 2018 16:46:56	S	Std
7.5 - 1	S7.5	28111	-	7.5000245.1	PPB	26 Apr 2018 16:49:03	S	Std
10 - 1	S10	35371	-	245.1	PPB	26 Apr 2018 17:01:08	S	Std
ICV18 - 1	ICV18	13720	3.8027	-245.1	PPB	26 Apr 2018 17:03:45	U	SMPL
ICB18 - 1	ICB18	-140	-0.0438	-245.1	PPB	26 Apr 2018 17:05:51	U	SMPL
CCV48 - 1	CCV48	17458	4.8401	-245.1	PPB	26 Apr 2018 17:07:58	U	SMPL
CCB48 - 1	CCB48	-185	-0.0563	-245.1	PPB	26 Apr 2018 17:10:05	U	SMPL
CRA - 1	CRA	195	0.0492	-245.1	PPB	26 Apr 2018 17:12:13	U	SMPL
HighStd - 1	HighStd	35537	9.8574	-245.1	PPB	26 Apr 2018 17:14:19	U	SMPL
ChkStd - 1	ChkStd	25138	6.9715	-245.1	PPB	26 Apr 2018 17:16:26	U	SMPL
PB108631BL - 1	PBW	-231	-0.0691	-245.1	PPB	26 Apr 2018 17:18:35	U	SMPL
PB108631BS - 1	LCSW	13620	3.7749	-245.1	PPB	26 Apr 2018 17:20:44	U	SMPL
J2571-01 - 1	042418-DBRI-E-D-B	-191	-0.0580	-245.1	PPB	26 Apr 2018 17:22:52	U	SMPL
J2571-02 - 1	042418-DBRI-E-D-M	-123	-0.0391	-245.1	PPB	26 Apr 2018 17:25:00	U	SMPL
J2571-03 - 1	042418-DBRI-E-D-S	-81	-0.0274	-245.1	PPB	26 Apr 2018 17:27:08	U	SMPL
J2571-04 - 1	042418-DBRI-E-U-B	-88	-0.0294	-245.1	PPB	26 Apr 2018 17:29:16	U	SMPL
J2571-05 - 1	042418-DBRI-E-U-M	-78	-0.0266	-245.1	PPB	26 Apr 2018 17:31:23	U	SMPL
CCV49 - 1	CCV49	17637	4.8897	-245.1	PPB	26 Apr 2018 17:33:29	U	SMPL
CCB49 - 1	CCB49	-143	-0.0446	-245.1	PPB	26 Apr 2018 17:35:36	U	SMPL
J2571-06 - 1	042418-DBRI-E-U-S	-116	-0.0372	-245.1	PPB	26 Apr 2018 17:37:43	U	SMPL
J2571-07 - 1	042418-SIDI-E-D-B	-95	-0.0313	-245.1	PPB	26 Apr 2018 17:39:50	U	SMPL
J2571-08 - 1	042418-SIDI-E-D-M	-89	-0.0297	-245.1	PPB	26 Apr 2018 17:41:57	U	SMPL
J2571-09 - 1	042418-SIDI-E-D-S	-84	-0.0283	-245.1	PPB	26 Apr 2018 17:44:04	U	SMPL
J2571-10 - 1	042418-SIDI-E-U-B	-81	-0.0274	-245.1	PPB	26 Apr 2018 17:46:11	U	SMPL
J2571-11 - 1	042418-SIDI-E-U-M	-79	-0.0269	-245.1	PPB	26 Apr 2018 17:48:18	U	SMPL
J2571-12 - 1	042418-SIDI-E-U-S	-98	-0.0322	-245.1	PPB	26 Apr 2018 17:50:26	U	SMPL
J2571-13 - 1	042418-JETB-E-D-M	-96	-0.0316	-245.1	PPB	26 Apr 2018 17:52:34	U	SMPL
J2571-14 - 1	042418-JETB-E-U-M	-97	-0.0319	-245.1	PPB	26 Apr 2018 17:54:42	U	SMPL
J2572-01 - 1	042518-SIDI-E-D-B	-70	-0.0244	-245.1	PPB	26 Apr 2018 17:56:49	U	SMPL
CCV50 - 1	CCV50	17363	4.8137	-245.1	PPB	26 Apr 2018 17:58:56	U	SMPL
CCB50 - 1	CCB50	-167	-0.0513	-245.1	PPB	26 Apr 2018 18:01:04	U	SMPL
J2572-02 - 1	042518-SIDI-E-D-M	-103	-0.0335	-245.1	PPB	26 Apr 2018 18:03:12	U	SMPL
J2572-03 - 1	042518-SIDI-E-D-S	-61	-0.0219	-245.1	PPB	26 Apr 2018 18:05:19	U	SMPL
J2572-04 - 1	042518-SIDI-E-U-B	-75	-0.0258	-245.1	PPB	26 Apr 2018 18:07:26	U	SMPL
J2572-05 - 1	042518-SIDI-E-U-M	-80	-0.0272	-245.1	PPB	26 Apr 2018 18:09:34	U	SMPL
J2572-06 - 1	042518-SIDI-E-U-S	-71	-0.0247	-245.1	PPB	26 Apr 2018 18:11:41	U	SMPL
J2572-06DUP - 1	042518-SIDI-E-U-SDUP	-57	-0.0208	-245.1	PPB	26 Apr 2018 18:13:49	U	SMPL
J2572-06MS - 1	042518-SIDI-E-U-SMS	15650	4.3383	-245.1	PPB	26 Apr 2018 18:15:56	U	SMPL
J2572-06MSD - 1	042518-SIDI-E-U-SMSD	14703	4.0755	-245.1	PPB	26 Apr 2018 18:18:04	U	SMPL
J2572-06LX5 - 1		-131	-0.0413	-245.1	PPB	26 Apr 2018 18:20:13	U	SMPL
J2572-06A - 1		15594	4.3228	-245.1	PPB	26 Apr 2018 18:22:20	U	SMPL
CCV51 - 1	CCV51	17388	4.8206	-245.1	PPB	26 Apr 2018 18:24:27	U	SMPL
CCB51 - 1	CCB51	-182	-0.0555	-245.1	PPB	26 Apr 2018 18:26:34	U	SMPL