


7471B

LB94953
INSTRUMENT ID: CV2

Linear

 μ Abs.:

43267



A= 0.0000e+000

B= 2.3016e-004

C= 1.6482e-002

Rho= 0.9999362

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	4	0.000	4				
0.2	0.200	0.226	0.026	909	0.0 %	909				
2.5	2.500	2.435	-0.065	10510	0.0 %	10510				
5.0	5.000	4.981	-0.019	21569	0.0 %	21569				
7.5	7.500	7.566	0.066	32799	0.0 %	32799				
10.0	10.000	9.975	-0.025	43267	0.0 %	43267				

LB94953 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Apr 2018 15:47:16

Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	7471B	PPB	27 Apr 2018 15:48:28	S	Std
0.2 - 1	50.2	909	-	0.2000	7471B	PPB	27 Apr 2018 15:50:45	S	Std
2.5 - 1	52.5	10510	-	2.5000	7471B	PPB	27 Apr 2018 15:53:03	S	Std
5.0 - 1	55	21569	-	5.0000	7471B	PPB	27 Apr 2018 15:55:20	S	Std
7.5 - 1	57.5	32799	-	7.5000	7471B	PPB	27 Apr 2018 15:57:37	S	Std
10.0 - 1	510	43267	-	10.0000	7471B	PPB	27 Apr 2018 15:59:54	S	Std
ICV20 - 1	ICV20	16432	3.7985	-	7471B	PPB	27 Apr 2018 16:02:50	U	SMPL
ICB20 - 1	ICB20	-192	-0.0277	-	7471B	PPB	27 Apr 2018 16:05:05	U	SMPL
CCV57 - 1	CCV57	22059	5.0936	-	7471B	PPB	27 Apr 2018 16:07:22	U	SMPL
CCB57 - 1	CCB57	-205	-0.0307	-	7471B	PPB	27 Apr 2018 16:09:38	U	SMPL
CRA - 1	CRA	856	0.2135	-	7471B	PPB	27 Apr 2018 16:11:54	U	SMPL
HighStd - 1	HighStd	43913	10.1236	-	7471B	PPB	27 Apr 2018 16:14:10	U	SMPL
ChkStd - 1	ChkStd	30665	7.0744	-	7471B	PPB	27 Apr 2018 16:16:26	U	SMPL
PB108658BL - 1	PBS	-248	-0.0406	-	7471B	PPB	27 Apr 2018 16:18:44	U	SMPL
PB108658BS - 1	LCSS	15380	3.5564	-	7471B	PPB	27 Apr 2018 16:21:02	U	SMPL
J2597-01 - 1	KS-RO-133	1017	0.2506	-	7471B	PPB	27 Apr 2018 16:23:19	U	SMPL
J2607-01 - 1	NJD001-74	1949	0.4651	-	7471B	PPB	27 Apr 2018 16:25:36	U	SMPL
J2608-01 - 1	CO-5-1	7475	1.7370	-	7471B	PPB	27 Apr 2018 16:27:54	U	SMPL
J2609-01 - 1	SB-01A	31452	7.2556	-	7471B	PPB	27 Apr 2018 16:30:12	U	SMPL
J2609-02 - 1	SB-01B	124904	28.7647	-	7471B	PPB	27 Apr 2018 16:32:28	U	SMPL
CCV58 - 1	CCV58	20781	4.7995	-	7471B	PPB	27 Apr 2018 16:34:46	U	SMPL
CCB58 - 1	CCB58	-250	-0.0411	-	7471B	PPB	27 Apr 2018 16:37:09	U	SMPL
J2609-03 - 1	SB-01C	17186	3.9721	-	7471B	PPB	27 Apr 2018 16:39:25	U	SMPL
J2609-04 - 1	SB-01D	281918	HIGH	-	7471B	PPB	27 Apr 2018 16:41:42	X	
J2609-05 - 1	SB-01E	1508	0.3636	-	7471B	PPB	27 Apr 2018 16:46:11	U	SMPL
J2613-01 - 1	FK-GF-02-018-A	2402	0.5693	-	7471B	PPB	27 Apr 2018 16:48:27	U	SMPL
J2613-03 - 1	FK-GF-02-018-B	1455	0.3514	-	7471B	PPB	27 Apr 2018 16:50:44	U	SMPL
J2613-03DUP - 1	FK-GF-02-018-BDUP	1456	0.3516	-	7471B	PPB	27 Apr 2018 16:55:41	U	SMPL
J2613-03MS - 1	FK-GF-02-018-BMS	18016	4.1631	-	7471B	PPB	27 Apr 2018 16:57:59	U	SMPL
J2613-03MSD - 1	FK-GF-02-018-BMSD	18083	4.1785	-	7471B	PPB	27 Apr 2018 17:00:16	U	SMPL
J2613-03LX5 - 1		157	0.0526	-	7471B	PPB	27 Apr 2018 17:02:34	U	SMPL
J2613-03A - 1		19927	4.6029	-	7471B	PPB	27 Apr 2018 17:04:50	U	SMPL
CCV59 - 1	CCV59	21977	5.0748	-	7471B	PPB	27 Apr 2018 17:07:06	U	SMPL
CCB59 - 1	CCB59	-182	-0.0254	-	7471B	PPB	27 Apr 2018 17:09:23	U	SMPL
J2609-02DLX5 - 1		23434	5.4101	-	7471B	PPB	27 Apr 2018 17:11:40	U	SMPL
J2609-04DLX10 - 1		35560	8.2011	-	7471B	PPB	27 Apr 2018 17:13:57	U	SMPL
CCV60 - 1	CCV60	21754	5.0234	-	7471B	PPB	27 Apr 2018 17:16:59	U	SMPL
CCB60 - 1	CCB60	-178	-0.0245	-	7471B	PPB	27 Apr 2018 17:19:15	U	SMPL