



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94958

Ambient Barometric Pressure (mmHg): 750

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	83.9	04/27/2018	09:02
2	J2564-07		1	>100.0	>212.0	>212.0	04/27/2018	09:40
3	J2564-07DUP		1	>100.0	>212.0	>212.0	04/27/2018	10:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

LB94958

WorkList Name : FP-042718

WorkList ID : 111348

Date : 4/27/2018 8:17:41 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/01/2018	Water	J2564-07	Flash Point	Cool 4 deg C	PSEG01		LS-UB-102	04/26/2018	1010A

Date/Time 04/27/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 04/27/18 10:15 AM
 Received by: CP
 Relinquished by: KS