



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94960

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2322
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	04/27/2018	16:45
2	CAL2	1	Water			21.2	7.00	04/27/2018	16:47
3	CAL3	1	Water			21.5	10.03	04/27/2018	16:49
4	ICV	1	Water			21.2	7.00	04/27/2018	16:51
5	CCV1	1	Water			21.4	2.00	04/27/2018	16:53
6	J2158-08	1	Solid	20.2	20	21.6	7.74	04/27/2018	16:55
7	J2158-08DUP	1	Solid	20.4	20	21.5	7.75	04/27/2018	16:57
8	J2165-08	1	Solid	20.4	20	21.4	7.63	04/27/2018	16:59
9	J2166-06	1	Solid	20.2	20	21.6	6.86	04/27/2018	17:01
10	J2615-03	1	Solid	20.5	20	21.5	8.48	04/27/2018	17:03
11	J2616-06	1	Solid	20.1	20	21.5	8.74	04/27/2018	17:05
12	J2616-12	1	Solid	20.3	20	21.9	8.68	04/27/2018	17:07
13	J2616-18	1	Solid	20.5	20	21.7	10.90	04/27/2018	17:09
14	J2616-24	1	Solid	20.2	20	21.7	11.79	04/27/2018	17:11
15	J2616-30	1	Solid	20.3	20	21.6	11.73	04/27/2018	17:13
16	CCV2	1	Water			21.5	12.00	04/27/2018	17:15
17	J2616-36	1	Solid	20.2	20	21.4	11.57	04/27/2018	17:17
18	J2616-46	1	Solid	20.3	20	21.5	9.08	04/27/2018	17:19
19	CCV3	1	Water			21.3	2.00	04/27/2018	17:21