


7470A

LB 94975
INSTRUMENT ID: CY1

Linear

 μ Abs.:

37531



A= 0.0000e+000

B= 2.6684e-004

C= 3.1723e-002

Rho= 0.9999360

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.036	0.036	17	0.000	17				
0.2	0.200	0.206	0.006	654	0.0 %	654				
2.5	2.500	2.503	0.003	9262	0.0 %	9262				
5.0	5.000	4.917	-0.083	18309	0.0 %	18309				
7.5	7.500	7.491	-0.009	27953	0.0 %	27953				
10.0	10.000	10.046	0.046	37531	0.0 %	37531				

LB94975 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 30 Apr 2018 11:35:05					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	17	-	0.0000	7470A	PPB	30 Apr 2018 11:38:51	S	Std
0.2 - 1	50.2	654	-	0.2000	7470A	PPB	30 Apr 2018 11:41:00	S	Std
2.5 - 1	52.5	9262	-	2.5000	7470A	PPB	30 Apr 2018 11:43:08	S	Std
5.0 - 1	55	18309	-	5.0000	7470A	PPB	30 Apr 2018 11:45:17	S	Std
7.5 - 1	57.5	27953	-	7.5000	7470A	PPB	30 Apr 2018 11:47:24	S	Std
10.0 - 1	510	37531	-	10.0000	7470A	PPB	30 Apr 2018 11:49:31	S	Std
ICV21 - 1	ICV21	14434	3.8833	-	7470A	PPB	30 Apr 2018 11:52:08	U	SMPL
ICB21 - 1	ICB21	-144	-0.0067	-	7470A	PPB	30 Apr 2018 11:54:15	U	SMPL
CCV61 - 1	CCV61	18768	5.0397	-	7470A	PPB	30 Apr 2018 11:56:21	U	SMPL
CCB61 - 1	CCB61	-148	-0.0078	-	7470A	PPB	30 Apr 2018 11:58:28	U	SMPL
CRA - 1	CRA	660	0.2078	-	7470A	PPB	30 Apr 2018 12:00:35	U	SMPL
HighStd - 1	HighStd	37682	10.0867	-	7470A	PPB	30 Apr 2018 12:08:56	U	SMPL
ChkStd - 1	ChkStd	25950	6.9562	-	7470A	PPB	30 Apr 2018 12:11:02	U	SMPL
PB108699BL - 1	PBW	-220	-0.0270	-	7470A	PPB	30 Apr 2018 12:13:11	U	SMPL
PB108699BS - 1	LCSW	13440	3.6180	-	7470A	PPB	30 Apr 2018 12:15:19	U	SMPL
PB108699TB - 1	PB108654TB	-112	0.0018	-	7470A	PPB	30 Apr 2018 12:17:27	U	SMPL
J2592-01 - 1	OD-1	-36	0.0221	-	7470A	PPB	30 Apr 2018 12:19:35	U	SMPL
J2592-02 - 1	OD-3-1	-7	0.0299	-	7470A	PPB	30 Apr 2018 12:21:43	U	SMPL
J2592-03 - 1	WASTE-OIL	11	0.0347	-	7470A	PPB	30 Apr 2018 12:23:51	U	SMPL
J2597-03 - 1	KS-RO-133	15	0.0357	-	7470A	PPB	30 Apr 2018 12:25:58	U	SMPL
CCV62 - 1	CCV62	18629	5.0026	-	7470A	PPB	30 Apr 2018 12:28:05	U	SMPL
CCB62 - 1	CCB62	-141	-0.0059	-	7470A	PPB	30 Apr 2018 12:30:12	U	SMPL
J2607-02 - 1	NJD001-74	-47	0.0192	-	7470A	PPB	30 Apr 2018 12:32:20	U	SMPL
J2608-02 - 1	CO-5-1	-27	0.0245	-	7470A	PPB	30 Apr 2018 12:34:26	U	SMPL
J2610-01 - 1	SB-01-WC	123	0.0645	-	7470A	PPB	30 Apr 2018 12:36:32	U	SMPL
J2610-01DUP - 1	SB-01-WCDUP	48	0.0445	-	7470A	PPB	30 Apr 2018 12:38:39	U	SMPL
J2610-01MS - 1	SB-01-WCMS	13038	3.5108	-	7470A	PPB	30 Apr 2018 12:45:20	U	SMPL
J2610-01MSD - 1	SB-01-WCMSD	12966	3.4915	-	7470A	PPB	30 Apr 2018 12:49:53	U	SMPL
PB108700BL - 1	PBW	-103	0.0042	-	7470A	PPB	30 Apr 2018 12:52:01	U	SMPL
PB108700BS - 1	LCSW	14244	3.8326	-	7470A	PPB	30 Apr 2018 12:56:33	U	SMPL
PB108700TB - 1	PB108656TB	-110	0.0024	-	7470A	PPB	30 Apr 2018 12:58:41	U	SMPL
J2564-07 - 1	LS-UB-102	-1	0.0315	-	7470A	PPB	30 Apr 2018 13:00:48	U	SMPL
CCV63 - 1	CCV63	18483	4.9637	-	7470A	PPB	30 Apr 2018 13:02:55	U	SMPL
CCB63 - 1	CCB63	-153	-0.0091	-	7470A	PPB	30 Apr 2018 13:05:02	U	SMPL
J2564-07DUP - 1	LS-UB-102DUP	-13	0.0283	-	7470A	PPB	30 Apr 2018 13:07:10	U	SMPL
J2564-07MS - 1	LS-UB-102MS	14342	3.8587	-	7470A	PPB	30 Apr 2018 13:09:17	U	SMPL
J2564-07MSD - 1	LS-UB-102MSD	13121	3.5329	-	7470A	PPB	30 Apr 2018 13:11:24	U	SMPL
J2610-01LX5 - 1		-92	0.0072	-	7470A	PPB	30 Apr 2018 13:13:31	U	SMPL
J2610-01A - 1		14359	3.8633	-	7470A	PPB	30 Apr 2018 13:15:39	U	SMPL
J2564-07LX5 - 1		-132	-0.0035	-	7470A	PPB	30 Apr 2018 13:17:46	U	SMPL
J2564-07A - 1		14289	3.8446	-	7470A	PPB	30 Apr 2018 13:19:55	U	SMPL
CCV64 - 1	CCV64	18725	5.0283	-	7470A	PPB	30 Apr 2018 13:22:03	U	SMPL
CCB64 - 1	CCB64	-165	-0.0123	-	7470A	PPB	30 Apr 2018 13:24:12	U	SMPL