

=====

Test results

Aquakem 7.2AQ1

LB94983
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS.

4/30/2018 12:21

Test: Total CNMg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.863	0.0	0.096	
ICB1	0.301	0.0	0.004	
CCV1	244.132	0.0	0.239	
CCB1	0.107	0.0	0.004	
PB108693BL	-0.224	0.0	0.003	
PB108693BS	96.174	0.0	0.097	
LOWPB108693	10.008	0.0	0.013	
HIGHPB108693	494.724	0.0	0.482	
J2604-05	1778.084	0.0	1.722	Test limit high
CCV2	253.017	0.0	0.248	
CCB2	0.391	0.0	0.004	
J2604-05DLX10	183.750	0.0	0.181	
CCV3	258.598	0.0	0.253	
CCB3	-0.569	0.0	0.003	

N	14
Mean	243.883
SD	465.4548
CV%	190.85

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 30 11:04:23 2018

Mon Apr 30 11:55:06 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.2938	µg/l	4/30/2018 11:04:23	
5.OPPBCN	A	Total CN	P	5.6698	µg/l	4/30/2018 11:04:24	
10PPBCN	A	Total CN	P	9.7088	µg/l	4/30/2018 11:04:25	
50PPBCN	A	Total CN	P	50.455	µg/l	4/30/2018 11:04:26	
100PPBCN	A	Total CN	P	100.2006	µg/l	4/30/2018 11:04:27	
250PPBCN	A	Total CN	P	248.6923	µg/l	4/30/2018 11:04:28	
500PPBCN	A	Total CN	P	500.5674	µg/l	4/30/2018 11:04:29	
ICV1	S	Total CN	P	95.8626	µg/l	4/30/2018 11:29:29	
ICB1	S	Total CN	P	0.3011	µg/l	4/30/2018 11:29:30	
CCV1	S	Total CN	P	244.1315	µg/l	4/30/2018 11:29:31	
CCB1	S	Total CN	P	0.1073	µg/l	4/30/2018 11:29:32	
PB108693BL	S	Total CN	P	-0.2239	µg/l	4/30/2018 11:29:33	
PB108693BS	S	Total CN	P	96.1739	µg/l	4/30/2018 11:29:34	
LOWPB108693	S	Total CN	P	10.0079	µg/l	4/30/2018 11:29:35	
HIGHPB108693	S	Total CN	P	494.7245	µg/l	4/30/2018 11:29:36	
J2604-05	S	Total CN	P	1778.084	µg/l	4/30/2018 11:29:38	
CCV2	S	Total CN	P	253.0174	µg/l	4/30/2018 11:33:48	
CCB2	S	Total CN	P	0.3909	µg/l	4/30/2018 11:33:49	
J2604-05DLX10	S	Total CN	P	183.7502	µg/l	4/30/2018 11:55:04	
CCV3	S	Total CN	P	258.5978	µg/l	4/30/2018 11:55:05	
CCB3	S	Total CN	P	-0.5687	µg/l	4/30/2018 11:55:06	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : MS

4/30/2018 11:04


Test Total CN

Accepted 4/30/2018 11:04

Factor 1035
Bias 0.004

Coeff. of det. 0.999986

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.2938	0.0000	
2	5.0PPBCN	0.009	5.6698	5.0000	
3	10PPBCN	0.013	9.7088	10.0000	
4	50PPBCN	0.052	50.4550	50.0000	
5	100PPBCN	0.100	100.2006	100.0000	
6	250PPBCN	0.244	248.6923	250.0000	
7	500PPBCN	0.487	500.5674	500.0000	