



### Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94998

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2322
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	04/30/2018	16:10
2	CAL2	1	Water			20.2	7.00	04/30/2018	16:12
3	CAL3	1	Water			20.4	10.03	04/30/2018	16:14
4	ICV	1	Water			20.5	7.00	04/30/2018	16:16
5	CCV1	1	Water			20.3	2.00	04/30/2018	16:18
6	J2157-14	1	Solid	20.2	20	20.3	8.08	04/30/2018	16:20
7	J2157-14DUP	1	Solid	20.3	20	20.4	8.09	04/30/2018	16:22
8	J2157-16	1	Solid	20.5	20	20.2	8.12	04/30/2018	16:24
9	J2157-18	1	Solid	20.2	20	20.2	8.37	04/30/2018	16:26
10	J2157-20	1	Solid	20.4	20	20.3	8.05	04/30/2018	16:28
11	J2157-22	1	Solid	20.3	20	20.2	7.83	04/30/2018	16:30
12	J2642-02	1	Solid	20.4	20	20.4	6.47	04/30/2018	16:32
13	J2646-02	1	Solid	20.2	20	20.1	5.04	04/30/2018	16:34
14	J2647-02	1	Solid	20.3	20	20.3	5.57	04/30/2018	16:36
15	CCV2	1	Water			20.5	12.00	04/30/2018	16:38

WORKLIST(Hardcopy Internal Chain)

LB 94998

Worklist Name : Corrosivity043018

Worklist ID : 111447

Date : 4/30/2018 3:53:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2018	Solid	J2157-14	Corrosivity	Cool 4 deg C	PSEG05		CL-02-043018-A	04/30/2018	9045C
04/30/2018	Solid	J2157-16	Corrosivity	Cool 4 deg C	PSEG05		CL-02-043018-B	04/30/2018	9045C
04/30/2018	Solid	J2157-18	Corrosivity	Cool 4 deg C	PSEG05		CL-02-043018-C	04/30/2018	9045C
04/30/2018	Solid	J2157-20	Corrosivity	Cool 4 deg C	PSEG05		CL-02-043018-D	04/30/2018	9045C
04/30/2018	Solid	J2157-22	Corrosivity	Cool 4 deg C	PSEG05		CL-02-043018-E	04/30/2018	9045C
04/27/2018	Solid	J2642-02	Corrosivity	Cool 4 deg C	VIAS01	H11	WL-PH3-ZEO-01	04/27/2018	9045C
04/30/2018	Solid	J2646-02	Corrosivity	Cool 4 deg C	PSEG02	B51	CS-DRUM-1-OILY-WIPE	04/30/2018	9045C
04/30/2018	Solid	J2647-02	Corrosivity	Cool 4 deg C	PSEG02	B51	CS-DRUM-2-OILY-WIPE	04/30/2018	9045C

Date/Time 04/30/18 4:05 PM

Received by: SP

Relinquished by: CP

Date/Time 04/30/18 4:45 PM

Received by: CP

Relinquished by: SP