


7470A

LB 9500 2
INSTRUMENT ID: CV2

Linear

 μ Abs.:

36499



A= 0.0000e+000

B= 2.7507e-004

C= 3.1689e-002

Rho= 0.9999042

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	-33	0.000	-33				
0.2	0.200	0.199	-0.001	607	0.0 %	607				
2.5	2.500	2.538	0.038	9112	0.0 %	9112				
5.0	5.000	4.938	-0.062	17835	0.0 %	17835				
7.5	7.500	7.432	-0.068	26902	0.0 %	26902				
10.0	10.000	10.071	0.071	36499	0.0 %	36499				

LB95002 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 May 2018 08:40:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-33	-	0.0000	7470A	PPB	01 May 2018 08:41:00	S	Std
0.2 - 1	50.2	607	-	0.2000	7470A	PPB	01 May 2018 08:43:09	S	Std
2.5 - 1	52.5	9112	-	2.5000	7470A	PPB	01 May 2018 08:45:18	S	Std
5.0 - 1	55	17835	-	5.0000	7470A	PPB	01 May 2018 08:47:26	S	Std
7.5 - 1	57.5	26902	-	7.5000	7470A	PPB	01 May 2018 08:49:34	S	Std
10.0 - 1	510	36499	-	10.0000	7470A	PPB	01 May 2018 08:54:25	S	Std
ICV23 - 1	ICV23	14142	3.9217	-	7470A	PPB	01 May 2018 08:56:53	U	SMPL
ICB23 - 1	ICB23	-212	-0.0266	-	7470A	PPB	01 May 2018 08:59:00	U	SMPL
CCV70 - 1	CCV70	18105	5.0118	-	7470A	PPB	01 May 2018 09:01:08	U	SMPL
CCB70 - 1	CCB70	-214	-0.0272	-	7470A	PPB	01 May 2018 09:03:14	U	SMPL
CRA - 1	CRA	574	0.1896	-	7470A	PPB	01 May 2018 09:05:22	U	SMPL
HighStd - 1	HighStd	36747	10.1397	-	7470A	PPB	01 May 2018 09:07:28	U	SMPL
ChkStd - 1	ChkStd	25451	7.0325	-	7470A	PPB	01 May 2018 09:09:35	U	SMPL
PB108732BL - 1	PBW	-294	-0.0492	-	7470A	PPB	01 May 2018 09:11:45	U	SMPL
PB108732BS - 1	LCSW	14464	4.0103	-	7470A	PPB	01 May 2018 09:16:28	U	SMPL
PB108732TB - 1	PB108703TB	-157	-0.0115	-	7470A	PPB	01 May 2018 09:18:35	U	SMPL
J2158-08 - 1	OK-02-042718	-66	0.0135	-	7470A	PPB	01 May 2018 09:20:43	U	SMPL
J2158-08DUP - 1	OK-02-042718DUP	-52	0.0174	-	7470A	PPB	01 May 2018 09:22:51	U	SMPL
J2158-08MS - 1	OK-02-042718MS	14304	3.9663	-	7470A	PPB	01 May 2018 09:29:28	U	SMPL
J2158-08MSD - 1	OK-02-042718MSD	14082	3.9052	-	7470A	PPB	01 May 2018 09:31:34	U	SMPL
CCV71 - 1	CCV71	18245	5.0503	-	7470A	PPB	01 May 2018 09:33:42	U	SMPL
CCB71 - 1	CCB71	-204	-0.0244	-	7470A	PPB	01 May 2018 09:35:49	U	SMPL
J2165-08 - 1	SU-04-042718	-48	0.0185	-	7470A	PPB	01 May 2018 09:37:58	U	SMPL
J2166-06 - 1	TR-05-042718	-55	0.0166	-	7470A	PPB	01 May 2018 09:40:04	U	SMPL
J2615-03 - 1	PLANTER-AREAS-PRECHAR	-23	0.0254	-	7470A	PPB	01 May 2018 09:42:11	U	SMPL
J2616-06 - 1	ROBEL-1	-15	0.0276	-	7470A	PPB	01 May 2018 09:44:18	U	SMPL
J2616-12 - 1	ROBEL-2	-41	0.0204	-	7470A	PPB	01 May 2018 09:46:26	U	SMPL
J2616-18 - 1	CONCRETE-1	-65	0.0138	-	7470A	PPB	01 May 2018 09:48:33	U	SMPL
J2616-24 - 1	CONCRETE-2	-54	0.0168	-	7470A	PPB	01 May 2018 09:50:41	U	SMPL
J2616-30 - 1	CONCRETE-3	-59	0.0155	-	7470A	PPB	01 May 2018 09:52:50	U	SMPL
J2616-36 - 1	CONCRETE-4	-49	0.0182	-	7470A	PPB	01 May 2018 09:54:58	U	SMPL
J2616-46 - 1	HOPE-CREEK-SOIL	-27	0.0243	-	7470A	PPB	01 May 2018 09:57:04	U	SMPL
CCV72 - 1	CCV72	18223	5.0443	-	7470A	PPB	01 May 2018 09:59:11	U	SMPL
CCB72 - 1	CCB72	-189	-0.0203	-	7470A	PPB	01 May 2018 10:01:18	U	SMPL
J2158-08LX5 - 1		-70	0.0124	-	7470A	PPB	01 May 2018 10:03:27	U	SMPL
J2158-08A - 1		14214	3.9415	-	7470A	PPB	01 May 2018 10:05:33	U	SMPL
CCV73 - 1	CCV73	18366	5.0836	-	7470A	PPB	01 May 2018 10:07:40	U	SMPL
CCB73 - 1	CCB73	-196	-0.0222	-	7470A	PPB	01 May 2018 10:09:48	U	SMPL