


245.1

LB95004
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35136



A= 0.0000e+000

B= 2.8195e-004

C= -3.9471e-003

Rho= 0.9998709

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-52	0.000	-52				
0.05	0.050	0.049	-0.001	187	0.0 %	187				
0.2	0.200	0.192	-0.008	695	0.0 %	695				
2.5	2.500	2.461	-0.039	8744	0.0 %	8744				
5.0	5.000	5.073	0.073	18005	0.0 %	18005				
7.5	7.500	7.591	0.091	26937	0.0 %	26937				
10.0	10.000	9.903	-0.097	35136	0.0 %	35136				

LB95004 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 May 2018 11:29:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-52	-	0.0000	245.1	PPB	01 May 2018 11:33:46	S	Std
0.05 - 1	30.05	187	-	0.0500	245.1	PPB	01 May 2018 11:35:54	S	Std
0.2 - 1	50.2	695	-	0.2000	245.1	PPB	01 May 2018 11:38:03	S	Std
2.5 - 1	52.5	8744	-	2.5000	245.1	PPB	01 May 2018 11:40:11	S	Std
5.0 - 1	55	18005	-	5.0000	245.1	PPB	01 May 2018 11:42:19	S	Std
7.5 - 1	57.5	26937	-	7.5000	245.1	PPB	01 May 2018 11:44:27	S	Std
10.0 - 1	510	35136	-	10.0000	245.1	PPB	01 May 2018 11:49:08	S	Std
ICV24 - 1	ICV24	14137	3.9820	-245.1	PPB	PPB	01 May 2018 11:54:12	U	SMPL
ICB24 - 1	ICB24	-198	-0.0598	-245.1	PPB	PPB	01 May 2018 11:56:17	U	SMPL
CCV74 - 1	CCV74	18243	5.1397	-245.1	PPB	PPB	01 May 2018 11:58:25	U	SMPL
CCB74 - 1	CCB74	-191	-0.0578	-245.1	PPB	PPB	01 May 2018 12:00:31	U	SMPL
CRA - 1	CRA	190	0.0496	-245.1	PPB	PPB	01 May 2018 12:02:39	U	SMPL
HighStd - 1	HighStd	35288	9.9456	-245.1	PPB	PPB	01 May 2018 12:04:45	U	SMPL
ChkStd - 1	ChkStd	26065	7.3452	-245.1	PPB	PPB	01 May 2018 12:06:52	U	SMPL
PB108731BL - 1	PBW	-212	-0.0637	-245.1	PPB	PPB	01 May 2018 12:09:02	U	SMPL
PB108731BS - 1	LCSW	13242	3.7297	-245.1	PPB	PPB	01 May 2018 12:11:12	U	SMPL
J2617-01 - 1	042618-DBRI-F-D-B	-184	-0.0558	-245.1	PPB	PPB	01 May 2018 12:13:19	U	SMPL
J2617-02 - 1	042618-DBRI-F-D-M	-77	-0.0257	-245.1	PPB	PPB	01 May 2018 12:15:28	U	SMPL
J2617-03 - 1	042618-DBRI-F-D-S	-98	-0.0316	-245.1	PPB	PPB	01 May 2018 12:17:36	U	SMPL
J2617-04 - 1	042618-DBRI-F-U-B	-90	-0.0293	-245.1	PPB	PPB	01 May 2018 12:19:45	U	SMPL
J2617-05 - 1	042618-DBRI-F-U-M	54	0.0113	-245.1	PPB	PPB	01 May 2018 12:21:51	U	SMPL
CCV75 - 1	CCV75	18256	5.1434	-245.1	PPB	PPB	01 May 2018 12:23:58	U	SMPL
CCB75 - 1	CCB75	-187	-0.0567	-245.1	PPB	PPB	01 May 2018 12:26:05	U	SMPL
J2617-06 - 1	042618-DBRI-F-U-S	-104	-0.0333	-245.1	PPB	PPB	01 May 2018 12:28:13	U	SMPL
J2617-07 - 1	042618-SIDI-E-D-B	-52	-0.0186	-245.1	PPB	PPB	01 May 2018 12:30:19	U	SMPL
J2617-08 - 1	042618-SIDI-E-D-M	-70	-0.0237	-245.1	PPB	PPB	01 May 2018 12:32:26	U	SMPL
J2617-09 - 1	042618-SIDI-E-D-S	-102	-0.0327	-245.1	PPB	PPB	01 May 2018 12:34:33	U	SMPL
J2617-10 - 1	042618-SIDI-E-U-B	-105	-0.0336	-245.1	PPB	PPB	01 May 2018 12:36:41	U	SMPL
J2617-11 - 1	042618-SIDI-E-U-M	-56	-0.0197	-245.1	PPB	PPB	01 May 2018 12:38:48	U	SMPL
J2617-12 - 1	042618-SIDI-E-U-S	-78	-0.0259	-245.1	PPB	PPB	01 May 2018 12:40:56	U	SMPL
J2617-13 - 1	042618-JETB-E-D-M	-50	-0.0180	-245.1	PPB	PPB	01 May 2018 12:43:05	U	SMPL
J2617-14 - 1	042618-JETB-E-U-M	-65	-0.0223	-245.1	PPB	PPB	01 May 2018 12:45:14	U	SMPL
J2617-14DUP - 1	042618-JETB-E-U-MDUP	-52	-0.0186	-245.1	PPB	PPB	01 May 2018 12:47:21	U	SMPL
CCV76 - 1	CCV76	17993	5.0692	-245.1	PPB	PPB	01 May 2018 12:49:28	U	SMPL
CCB76 - 1	CCB76	-184	-0.0558	-245.1	PPB	PPB	01 May 2018 12:51:35	U	SMPL
J2617-14MS - 1	042618-JETB-E-U-MMS	15655	4.4100	-245.1	PPB	PPB	01 May 2018 12:53:44	U	SMPL
J2617-14MSD - 1	042618-JETB-E-U-MMSD	15451	4.3525	-245.1	PPB	PPB	01 May 2018 12:55:50	U	SMPL
J2617-14LX5 - 1		-172	-0.0524	-245.1	PPB	PPB	01 May 2018 12:57:57	U	SMPL
J2617-14A - 1		15969	4.4986	-245.1	PPB	PPB	01 May 2018 13:00:05	U	SMPL
CCV77 - 1	CCV77	17879	5.0371	-245.1	PPB	PPB	01 May 2018 13:02:13	U	SMPL
CCB77 - 1	CCB77	-215	-0.0646	-245.1	PPB	PPB	01 May 2018 13:04:20	U	SMPL