


ISM02.4

LB 95017
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41532



A= 0.0000e+000

B= 2.4272e-004

C= -5.5557e-002

Rho= 0.9999200

Accept=Accepted

Slope
 $y = 1.37125 \times 10^{-4}x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	0.016	0.016	293	0.000	293					0
0.05	0.050	0.045	-0.005	415	0.0 %	415					-10
0.2	0.200	0.179	-0.021	967	0.0 %	967					-11
2.5	2.500	2.489	-0.011	10484	0.0 %	10484					0
5.0	5.000	5.082	0.082	21165	0.0 %	21165					2
7.5	7.500	7.414	-0.086	30776	0.0 %	30776					-1
10.0	10.000	10.025	0.025	41532	0.0 %	41532					0

ABS
51118

LB95017
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	293	-		0 ISM02.4	PPB	5/1/2018 15:12	Std
	0.05 S0.05	415	-		0.05 ISM02.4	PPB	5/1/2018 15:14	Std
	0.2 S0.2	967	-		0.2 ISM02.4	PPB	5/1/2018 15:16	Std
	2.5 S2.5	10484	-		2.5 ISM02.4	PPB	5/1/2018 15:19	Std
	5 S5	21165	-		5 ISM02.4	PPB	5/1/2018 15:21	Std
	7.5 S7.5	30776	-		7.5 ISM02.4	PPB	5/1/2018 15:25	Std
	10 S10	41532	-		10 ISM02.4	PPB	5/1/2018 15:28	Std
ICV	ICV	16001	3.8282	-	ISM02.4	PPB	5/1/2018 15:31	SMPL
ICB	ICB	130	-0.024	-	ISM02.4	PPB	5/1/2018 15:33	SMPL
CCV078	CCV078	21192	5.0881	-	ISM02.4	PPB	5/1/2018 15:35	SMPL
CCB078	CCB078	121	-0.0262	-	ISM02.4	PPB	5/1/2018 15:38	SMPL
PB108753BL	PBW003	61	-0.0408	-	ISM02.4	PPB	5/1/2018 15:40	SMPL
J2516-01	MYAHG2	10	-0.0531	-	ISM02.4	PPB	5/1/2018 15:42	SMPL
J2516-02	MYAHG4	34	-0.0473	-	ISM02.4	PPB	5/1/2018 15:45	SMPL
J2516-03	MYAHH8	47	-0.0441	-	ISM02.4	PPB	5/1/2018 15:47	SMPL
J2516-04	MYAHH8D	21	-0.0505	-	ISM02.4	PPB	5/1/2018 15:49	SMPL
J2516-05	MYAHH8S	4360	1.0027	-	ISM02.4	PPB	5/1/2018 15:51	SMPL
J2516-06	MYAHJ0	7463	1.7559	-	ISM02.4	PPB	5/1/2018 15:54	SMPL
J2516-07	MYAHK0	238	0.0022	-	ISM02.4	PPB	5/1/2018 15:56	SMPL
J2516-08	MYAHM0	54460	13.1629	-	ISM02.4	PPB	5/1/2018 15:58	SMPL
J2516-09	MYAHN0	350	0.0294	-	ISM02.4	PPB	5/1/2018 16:03	SMPL
J2516-10	MYAHN2	527	0.0724	-	ISM02.4	PPB	5/1/2018 16:06	SMPL
PB108754BL	PBW003	363	0.0326	-	ISM02.4	PPB	5/1/2018 16:08	SMPL
J2517-01	MYAHG1	848	0.1503	-	ISM02.4	PPB	5/1/2018 16:10	SMPL
J2517-02	MYAHG3	640	0.0998	-	ISM02.4	PPB	5/1/2018 16:12	SMPL
J2517-03	MYAHG5	1223	0.2413	-	ISM02.4	PPB	5/1/2018 16:15	SMPL
J2517-04	MYAHG7	414	0.0449	-	ISM02.4	PPB	5/1/2018 16:17	SMPL
J2517-05	MYAHH3	29669	7.1457	-	ISM02.4	PPB	5/1/2018 16:19	SMPL
J2517-06	MYAHH5	382	0.0372	-	ISM02.4	PPB	5/1/2018 16:22	SMPL
J2517-07	MYAHH7	622	0.0954	-	ISM02.4	PPB	5/1/2018 16:24	SMPL
J2517-08	MYAHH7D	629	0.0971	-	ISM02.4	PPB	5/1/2018 16:26	SMPL
CCV079	CCV079	20911	5.0199	-	ISM02.4	PPB	5/1/2018 16:28	SMPL
CCB079	CCB079	68	-0.0391	-	ISM02.4	PPB	5/1/2018 16:31	SMPL
J2517-09	MYAHH7S	4470	1.0294	-	ISM02.4	PPB	5/1/2018 16:33	SMPL
J2517-10	MYAHH9	83156	20.128	-	ISM02.4	PPB	5/1/2018 16:35	SMPL
J2517-11	MYAHJ1	91	-0.0335	-	ISM02.4	PPB	5/1/2018 16:40	SMPL
J2517-12	MYAHJ9	15208	3.6357	-	ISM02.4	PPB	5/1/2018 16:43	SMPL
J2517-13	MYAHK5	1325191	HIGH	-	ISM02.4	PPB	5/1/2018 16:45	
J2517-14	MYAHL1	37807	9.1209	-	ISM02.4	PPB	5/1/2018 16:50	SMPL
J2517-15	MYAHL9	465176	HIGH	-	ISM02.4	PPB	5/1/2018 16:52	
J2517-16	MYAHM1	414	0.0449	-	ISM02.4	PPB	5/1/2018 16:54	SMPL
J2517-17	MYAHM3	21	-0.0505	-	ISM02.4	PPB	5/1/2018 17:00	SMPL
J2517-18	MYAHM9	2002	0.4304	-	ISM02.4	PPB	5/1/2018 17:02	SMPL
J2517-19	MYAHN1	1605	0.334	-	ISM02.4	PPB	5/1/2018 17:04	SMPL
J2517-20	MYAHM7	701	0.1146	-	ISM02.4	PPB	5/1/2018 17:06	SMPL
J2517-21	MYAHP1	51530	12.4518	-	ISM02.4	PPB	5/1/2018 17:09	SMPL

LB95017

INSTRUMENT ID : CV1

CCV080	CCV080	20514	4.9236 -	ISM02.4	PPB	5/1/2018 17:11	SMPL
CCB080	CCB080	6	-0.0541 -	ISM02.4	PPB	5/1/2018 17:13	SMPL
J2516-08DLX2	MYAHM0	26993	6.4962 -	ISM02.4	PPB	5/1/2018 17:16	SMPL
J2517-10DLX5	MYAHH9	15240	3.6435 -	ISM02.4	PPB	5/1/2018 17:18	SMPL
J2517-15DLX10	MYAHL9	68984	16.6882 -	ISM02.4	PPB	5/1/2018 17:22	SMPL
J2517-21DLX2	MYAHP1	24326	5.8488 -	ISM02.4	PPB	5/1/2018 17:25	SMPL
J2517-13DLX50	MYAHK5	310842	75.3917 -	ISM02.4	PPB	5/1/2018 17:27	SMPL
J2517-15DL2X50	MYAHL9	13695	3.2685 -	ISM02.4	PPB	5/1/2018 17:30	SMPL
J2517-13DL2X500	MYAHK5	35362	8.5275 -	ISM02.4	PPB	5/1/2018 17:33	SMPL
CCV081	CCV081	22637	5.4389 -	ISM02.4	PPB	5/1/2018 17:35	SMPL
CCB081	CCB081	-9	-0.0577 -	ISM02.4	PPB	5/1/2018 17:37	SMPL