



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95020

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	05/01/2018	15:30
2	CAL2	1	Water			20.8	7.00	05/01/2018	15:32
3	CAL3	1	Water			21.0	10.03	05/01/2018	15:34
4	ICV	1	Water			21.3	7.00	05/01/2018	15:36
5	CCV1	1	Water			21.2	2.00	05/01/2018	15:38
6	J2667-11	1	Water			19.2	7.22	05/01/2018	15:40
7	J2667-11DUP	1	Water			19.2	7.23	05/01/2018	15:42
8	CCV2	1	Water			21.3	12.00	05/01/2018	15:44

WORKLIST(Hardcopy Internal Chain)

LB 95020

Worklist Name : PHJ2667

Worklist ID : 111495

Date : 5/1/2018 3:18:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/01/2018	Water	J2667-11	pH	Cool 4 deg C	PSEG01	B63	GROUNDWATER	05/01/2018	9040C

Date/Time

05/01/18 3:25 PM

Received by:

SP

Relinquished by:

CP

Date/Time

05/01/18 4:10 PM

Received by:

CP

Relinquished by:

SP