


7471B

LB 95025
INSTRUMENT ID: CV2

Linear

 μ Abs.:

44363



A= 0.0000e+000

B= 2.2215e-004

C= -1.0260e-002

Rho= 0.9995537

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	40	0.000	40				
0.2	0.200	0.204	0.004	963	0.0 %	963				
2.5	2.500	2.402	-0.098	10858	0.0 %	10858				
5.0	5.000	5.064	0.064	22841	0.0 %	22841				
7.5	7.500	7.687	0.187	34651	0.0 %	34651				
10.0	10.000	9.845	-0.155	44363	0.0 %	44363				

LB95025 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 May 2018 18:36:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.00007471B	PPB		01 May 2018 18:37:09	S	Std
0.2 - 1	50.2	963	-	0.20007471B	PPB		01 May 2018 18:39:27	S	Std
2.5 - 1	52.5	10858	-	2.50007471B	PPB		01 May 2018 18:41:44	S	Std
5.0 - 1	55	22841	-	5.00007471B	PPB		01 May 2018 18:44:02	S	Std
7.5 - 1	57.5	34651	-	7.50007471B	PPB		01 May 2018 18:46:19	S	Std
10.0 - 1	510	44363	-	10.00007471B	PPB		01 May 2018 18:48:36	S	Std
ICV25 - 1	ICV25	16516	3.6587	-7471B	PPB		01 May 2018 18:51:24	U	SMPL
ICB25 - 1	ICB25	-119	-0.0367	-7471B	PPB		01 May 2018 18:53:40	U	SMPL
CCV82 - 1	CCV82	22562	5.0018	-7471B	PPB		01 May 2018 18:55:57	U	SMPL
CCB82 - 1	CCB82	-103	-0.0331	-7471B	PPB		01 May 2018 18:58:12	U	SMPL
CRA - 1	CRA	967	0.2046	-7471B	PPB		01 May 2018 19:00:30	U	SMPL
HighStd - 1	HighStd	44306	9.8321	-7471B	PPB		01 May 2018 19:02:46	U	SMPL
ChkStd - 1	ChkStd	32056	7.1109	-7471B	PPB		01 May 2018 19:05:02	U	SMPL
PB108758BL - 1	PBS	-231	-0.0616	-7471B	PPB		01 May 2018 19:07:20	U	SMPL
PB108758BS - 1	LCSS	14607	3.2346	-7471B	PPB		01 May 2018 19:09:39	U	SMPL
J2157-13 - 1	CL-02-043018-A	2148	0.4669	-7471B	PPB		01 May 2018 19:19:49	U	SMPL
J2157-13DUP - 1	CL-02-043018-ADUP	2038	0.4425	-7471B	PPB		01 May 2018 19:22:06	U	SMPL
J2157-13MS - 1	CL-02-043018-AMS	18010	3.9906	-7471B	PPB		01 May 2018 19:24:24	U	SMPL
J2157-13MSD - 1	CL-02-043018-AMSD	18645	4.1317	-7471B	PPB		01 May 2018 19:26:42	U	SMPL
J2157-15 - 1	CL-02-043018-B	2032	0.4411	-7471B	PPB		01 May 2018 19:28:58	U	SMPL
CCV83 - 1	CCV83	22659	5.0233	-7471B	PPB		01 May 2018 19:31:14	U	SMPL
CCB83 - 1	CCB83	-116	-0.0360	-7471B	PPB		01 May 2018 19:33:30	U	SMPL
J2157-17 - 1	CL-02-043018-C	1917	0.4156	-7471B	PPB		01 May 2018 19:35:47	U	SMPL
J2157-19 - 1	CL-02-043018-D	4253	0.9345	-7471B	PPB		01 May 2018 19:38:03	U	SMPL
J2157-21 - 1	CL-02-043018-E	3974	0.8725	-7471B	PPB		01 May 2018 19:40:19	U	SMPL
J2642-01 - 1	WL-PH3-ZEO-01	1861	0.4032	-7471B	PPB		01 May 2018 19:42:36	U	SMPL
J2646-01 - 1	CS-DRUM-1-OILY-WIPE	3007	0.6577	-7471B	PPB		01 May 2018 19:44:52	U	SMPL
J2647-01 - 1	CS-DRUM-2-OILY-WIPE	120	0.0164	-7471B	PPB		01 May 2018 19:47:10	U	SMPL
J2650-01 - 1	PAINT-CHIPS	185685	41.2389	-7471B	PPB		01 May 2018 19:49:27	U	SMPL
J2650-03 - 1	PAINT-RESIDUE	43318	9.6127	-7471B	PPB		01 May 2018 19:51:45	U	SMPL
J2157-13LX5 - 1		338	0.0648	-7471B	PPB		01 May 2018 19:57:10	U	SMPL
J2157-13A - 1		20045	4.4427	-7471B	PPB		01 May 2018 19:59:26	U	SMPL
CCV84 - 1	CCV84	22509	4.9900	-7471B	PPB		01 May 2018 20:01:42	U	SMPL
CCB84 - 1	CCB84	-110	-0.0347	-7471B	PPB		01 May 2018 20:03:59	U	SMPL
J2650-01DLX10 - 1		20288	4.4966	-7471B	PPB		01 May 2018 20:06:16	U	SMPL
CCV85 - 1	CCV85	22767	5.0473	-7471B	PPB		01 May 2018 20:08:33	U	SMPL
CCB85 - 1	CCB85	-146	-0.0427	-7471B	PPB		01 May 2018 20:10:50	U	SMPL