



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95032

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	05/02/2018	09:10
2	J2667-15		1	>100.0	>212.0	>212.0	05/02/2018	09:35
3	J2667-15DUP		1	>100.0	>212.0	>212.0	05/02/2018	10:02

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-050218

WorkList ID : 111530

Date : 5/2/2018 8:18:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/04/2018	Water	J2667-15	B Flash Point	Cool 4 deg C	PSEG01	B63	GROUNDWATER	05/01/2018	1010A

Date/Time 05/02/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 05/02/18 10:10 AM
 Received by: CP
 Relinquished by: KS