

LB95034


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

36079



A= 0.0000e+000

B= 2.7384e-004

C= -3.5774e-003

Rho= 0.9995678

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	1	0.000	1				
0.2	0.200	0.198	-0.002	735	0.0 %	735				
2.5	2.500	2.508	0.008	9170	0.0 %	9170				
5.0	5.000	4.906	-0.094	17929	0.0 %	17929				
7.5	7.500	7.716	0.216	28188	0.0 %	28188				
10.0	10.000	9.876	-0.124	36079	0.0 %	36079				

LB95034 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 May 2018 12:21:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.0000	7470A	PPB	02 May 2018 12:22:03	S	Std
0.2 - 1	50.2	735	-	0.2000	7470A	PPB	02 May 2018 12:24:11	S	Std
2.5 - 1	52.5	9170	-	2.5000	7470A	PPB	02 May 2018 12:26:19	S	Std
5.0 - 1	55	17929	-	5.0000	7470A	PPB	02 May 2018 12:28:27	S	Std
7.5 - 1	57.5	28188	-	7.5000	7470A	PPB	02 May 2018 12:30:35	S	Std
10.0 - 1	510	36079	-	10.0000	7470A	PPB	02 May 2018 12:35:16	S	Std
ICV26 - 1	ICV26	14325	3.9192	-	7470A	PPB	02 May 2018 12:37:57	U	SMPL
ICB26 - 1	ICB26	-159	-0.0471	-	7470A	PPB	02 May 2018 12:40:03	U	SMPL
CCV86 - 1	CCV86	18273	5.0004	-	7470A	PPB	02 May 2018 12:42:11	U	SMPL
CCB86 - 1	CCB86	-149	-0.0444	-	7470A	PPB	02 May 2018 12:44:18	U	SMPL
CRA - 1	CRA	681	0.1829	-	7470A	PPB	02 May 2018 12:46:26	U	SMPL
HighStd - 1	HighStd	35767	9.7910	-	7470A	PPB	02 May 2018 12:48:32	U	SMPL
ChkStd - 1	ChkStd	25562	6.9964	-	7470A	PPB	02 May 2018 12:50:39	U	SMPL
PB108785BL - 1	PBW	-244	-0.0704	-	7470A	PPB	02 May 2018 12:52:49	U	SMPL
PB108785BS - 1	LCSW	13822	3.7815	-	7470A	PPB	02 May 2018 12:54:57	U	SMPL
PB108785TB - 1	PB108751TB	-110	-0.0337	-	7470A	PPB	02 May 2018 12:57:05	U	SMPL
J2157-14 - 1	CL-02-043018-A	-6	-0.0052	-	7470A	PPB	02 May 2018 12:59:13	U	SMPL
J2157-14DUP - 1	CL-02-043018-ADUP	0	-0.0036	-	7470A	PPB	02 May 2018 13:01:21	U	SMPL
J2157-14MS - 1	CL-02-043018-AMS	13841	3.7867	-	7470A	PPB	02 May 2018 13:06:21	U	SMPL
J2157-14MSD - 1	CL-02-043018-AMSD	13771	3.7675	-	7470A	PPB	02 May 2018 13:08:27	U	SMPL
CCV87 - 1	CCV87	18027	4.9330	-	7470A	PPB	02 May 2018 13:10:34	U	SMPL
CCB87 - 1	CCB87	-138	-0.0414	-	7470A	PPB	02 May 2018 13:12:40	U	SMPL
J2157-16 - 1	CL-02-043018-B	-3	-0.0044	-	7470A	PPB	02 May 2018 13:14:48	U	SMPL
J2157-18 - 1	CL-02-043018-C	5	-0.0022	-	7470A	PPB	02 May 2018 13:16:54	U	SMPL
J2157-20 - 1	CL-02-043018-D	20	0.0019	-	7470A	PPB	02 May 2018 13:19:01	U	SMPL
J2157-22 - 1	CL-02-043018-E	7	-0.0017	-	7470A	PPB	02 May 2018 13:21:08	U	SMPL
J2642-02 - 1	WL-PH3-ZEO-01	8	-0.0014	-	7470A	PPB	02 May 2018 13:23:15	U	SMPL
J2646-02 - 1	CS-DRUM-1-OILY-WIPE	30	0.0046	-	7470A	PPB	02 May 2018 13:25:23	U	SMPL
J2647-02 - 1	CS-DRUM-2-OILY-WIPE	-15	-0.0077	-	7470A	PPB	02 May 2018 13:27:31	U	SMPL
J2650-02 - 1	PAINT-CHIPS	6225	1.7011	-	7470A	PPB	02 May 2018 13:29:39	U	SMPL
J2650-04 - 1	PAINT-RESIDUE	433	0.1150	-	7470A	PPB	02 May 2018 13:31:47	U	SMPL
CCV88 - 1	CCV88	18112	4.9563	-	7470A	PPB	02 May 2018 13:33:54	U	SMPL
CCB88 - 1	CCB88	-151	-0.0449	-	7470A	PPB	02 May 2018 13:36:01	U	SMPL
J2157-14LX5 - 1		-18	-0.0085	-	7470A	PPB	02 May 2018 13:38:09	U	SMPL
J2157-14A - 1		13917	3.8075	-	7470A	PPB	02 May 2018 13:40:16	U	SMPL
CCV89 - 1	CCV89	17978	4.9196	-	7470A	PPB	02 May 2018 13:42:23	U	SMPL
CCB89 - 1	CCB89	-143	-0.0427	-	7470A	PPB	02 May 2018 13:44:30	U	SMPL