

2B 95041

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AS

5/2/2018 13:48

Test: CNEPA-NEW ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV ICV	95.408	0.0	0.110	
ICB ICB	-0.389	0.0	0.007	
CCV001 CCV001	238.931	0.0	0.265	
CCB001 CCB001	-0.252	0.0	0.007	
PB108695BL PBS006	-1.032	0.0	0.006	
J2606-22 MX2C82	41.558	0.0	0.052	
J2606-23 MX2C82D	40.578	0.0	0.051	
J2606-24 MX2C82S	133.634	0.0	0.152	
PB108696BL PBW005	-0.327	0.0	0.007	
J2606-10 MX2C78	282.819	0.0	0.312	
J2606-11 MX2C78D	283.969	0.0	0.314	
J2606-12 MX2C78S	384.090	0.0	0.422	
CCV002 CCV002	256.730	0.0	0.284	
CCB002 CCB002	0.042	0.0	0.008	

N	14
Mean	125.411
SD	136.2384
CV%	108.63

AS
05/02/18
AS
05/02/18

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed May 02 13:14:15 2018

Wed May 02 13:47:06 2018

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
S0.0	A	CNEPA-NEW	P	-0.9768	µg/l	5/2/2018 13:14:15
S5.0	A	CNEPA-NEW	P	4.2456	µg/l	5/2/2018 13:14:16
S10.0	A	CNEPA-NEW	P	9.2554	µg/l	5/2/2018 13:14:17
S100.0	A	CNEPA-NEW	P	100.2031	µg/l	5/2/2018 13:14:18
S250.0	A	CNEPA-NEW	P	254.5818	µg/l	5/2/2018 13:14:19
S500.0	A	CNEPA-NEW	P	497.6909	µg/l	5/2/2018 13:14:20
ICV ICV	S	CNEPA-NEW	P	95.4081	µg/l	5/2/2018 13:42:44
ICB ICB	S	CNEPA-NEW	P	-0.389	µg/l	5/2/2018 13:42:45
CCV001 CCV001	S	CNEPA-NEW	P	238.931	µg/l	5/2/2018 13:42:46
CCB001 CCB001	S	CNEPA-NEW	P	-0.2519	µg/l	5/2/2018 13:42:47
PB108695BL PBS006	S	CNEPA-NEW	P	-1.0318	µg/l	5/2/2018 13:42:48
J2606-22 MX2C82	S	CNEPA-NEW	P	41.5584	µg/l	5/2/2018 13:42:49
J2606-23 MX2C82D	S	CNEPA-NEW	P	40.5777	µg/l	5/2/2018 13:42:50
J2606-24 MX2C82S	S	CNEPA-NEW	P	133.6339	µg/l	5/2/2018 13:42:51
PB108696BL PBW005	S	CNEPA-NEW	P	-0.3266	µg/l	5/2/2018 13:42:52
J2606-10 MX2C78	S	CNEPA-NEW	P	282.819	µg/l	5/2/2018 13:42:53
J2606-11 MX2C78D	S	CNEPA-NEW	P	283.9688	µg/l	5/2/2018 13:42:54
J2606-12 MX2C78S	S	CNEPA-NEW	P	384.0897	µg/l	5/2/2018 13:47:03
CCV002 CCV002	S	CNEPA-NEW	P	256.7302	µg/l	5/2/2018 13:47:04
CCB002 CCB002	S	CNEPA-NEW	P	0.0424	µg/l	5/2/2018 13:47:05

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/2/2018 13:15


Test CNEPA-NEW

Accepted 5/2/2018 13:15

Factor 927.9
Bias 0.008

Coeff. of det. 0.999856

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.007	-0.9768	0.0000	
2	5.0PPBCN	0.012	4.2456	5.0000	
3	10PPBCN	0.018	9.2554	10.0000	
4	100PPBCN	0.116	100.2031	100.0000	
5	250PPBCN	0.282	254.5818	250.0000	
6	500PPBCN	0.544	497.6909	500.0000	