



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95054

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	05/02/2018	16:55
2	CAL2	1	Water			22.8	7.00	05/02/2018	16:57
3	CAL3	1	Water			22.6	10.03	05/02/2018	16:59
4	ICV	1	Water			22.8	7.00	05/02/2018	17:01
5	CCV1	1	Water			22.3	2.00	05/02/2018	17:03
6	J2703-04	1	Water			22.7	6.43	05/02/2018	17:05
7	J2703-04DUP	1	Water			22.8	6.44	05/02/2018	17:07
8	CCV2	1	Water			22.5	12.00	05/02/2018	17:09

WORKLIST(Hardcopy Internal Chain)

LB 95054

Worklist Name : corrosivity J2703

Worklist ID : 111563

Date : 5/2/2018 4:43:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2018	Water	J2703-04	Corrosivity	Cool 4 deg C	PSEG01	H62	MANHOLE	05/02/2018	9040C

Date/Time 05/02/18 4:50 PM

Received by: SP

Relinquished by: CP

Date/Time 05/02/18 5:15 PM

Received by: CP

Relinquished by: SP