


ISM02.4

LB95058
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42150



A= 0.0000e+000

B= 2.3807e-004

C= -3.8430e-002

Rho= 0.9999955

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.016	-0.016	96	0.000	96					
0.2	0.200	0.218	0.018	1077	0.0 %	1077					
2.5	2.500	2.491	-0.009	10626	0.0 %	10626					
5.0	5.000	5.008	0.008	21198	0.0 %	21198					
7.5	7.500	7.502	0.002	31675	0.0 %	31675					
10.0	10.000	9.996	-0.004	42150	0.0 %	42150					

000091

dB
512118

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	96	-		0 ISM02.4	PPB	5/2/2018 19:26	Std
	0.2 S0.2	1077	-		0.2 ISM02.4	PPB	5/2/2018 19:29	Std
	2.5 S2.5	10626	-		2.5 ISM02.4	PPB	5/2/2018 19:31	Std
	5 S5	21198	-		5 ISM02.4	PPB	5/2/2018 19:33	Std
	7.5 S7.5	31675	-		7.5 ISM02.4	PPB	5/2/2018 19:36	Std
	10 S10	42150	-		10 ISM02.4	PPB	5/2/2018 19:38	Std
ICV	ICV	16159	3.8085 -		ISM02.4	PPB	5/2/2018 19:41	SMPL
ICB	ICB	-59	-0.0525 -		ISM02.4	PPB	5/2/2018 19:43	SMPL
CCV096	CCV096	21277	5.0269 -		ISM02.4	PPB	5/2/2018 19:45	SMPL
CCB096	CCB096	-70	-0.0551 -		ISM02.4	PPB	5/2/2018 19:47	SMPL
PB108846BL	PBW003	100	-0.0146 -		ISM02.4	PPB	5/2/2018 19:50	SMPL
J2606-07	MX2C77	22741	5.3754 -		ISM02.4	PPB	5/2/2018 19:52	SMPL
J2606-08	MX2C77D	22736	5.3742 -		ISM02.4	PPB	5/2/2018 19:54	SMPL
J2606-09	MX2C77S	26525	6.2762 -		ISM02.4	PPB	5/2/2018 19:57	SMPL
CCV097	CCV097	21120	4.9895 -		ISM02.4	PPB	5/2/2018 20:01	SMPL
CCB097	CCB097	-98	-0.0618 -		ISM02.4	PPB	5/2/2018 20:03	SMPL