



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95063

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	05/02/2018	16:55
2	CAL2	1	Water			22.8	7.00	05/02/2018	16:57
3	CAL3	1	Water			22.6	10.03	05/02/2018	16:59
4	ICV	1	Water			22.8	7.00	05/02/2018	17:01
5	CCV1	1	Water			22.3	2.00	05/02/2018	17:03
6	J2604-02	1	Solid	20.02	20	21.4	7.12	05/02/2018	17:20
7	J2670-02	1	Solid	20.04	20	25.3	8.85	05/02/2018	17:22
8	J2670-02DUP	1	Solid	20.02	20	25.2	8.86	05/02/2018	17:24
9	J2679-02	1	Solid	20.03	20	25.2	8.29	05/02/2018	17:26
10	J2679-04	1	Solid	20.01	20	25.1	8.06	05/02/2018	17:28
11	J2679-06	1	Solid	20.04	20	25.2	8.05	05/02/2018	17:30
12	J2681-02	1	Solid	20.03	20	25.4	7.30	05/02/2018	17:32
13	J2681-04	1	Solid	20.03	20	25.3	7.52	05/02/2018	17:34
14	J2686-02	1	Solid	20.05	20	24.0	9.51	05/02/2018	17:36
15	J2686-04	1	Solid	20.02	20	24.5	6.65	05/02/2018	17:38
16	CCV2	1	Water			22.5	12.00	05/02/2018	17:40
17	J2695-02	1	Solid	20.03	20	24.2	10.22	05/02/2018	17:42
18	J2703-02	1	Solid	20.04	20	23.9	7.57	05/02/2018	17:44
19	J2703-02DUP	1	Solid	20.03	20	23.8	7.58	05/02/2018	17:46
20	J2706-04	1	Solid	20.04	20	24.8	9.19	05/02/2018	17:48

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J2706-05	1	Solid	20.02	20	24.9	10.40	05/02/2018	17:50
22	J2706-06	1	Solid	20.01	20	24.7	10.81	05/02/2018	17:52
23	J2710-03	1	Solid	20.03	20	26.0	8.42	05/02/2018	17:54
24	J2711-03	1	Solid	20.04	20	26.0	7.04	05/02/2018	17:56
25	J2712-03	1	Solid	20.01	20	26.2	7.59	05/02/2018	17:58
26	J2713-04	1	Solid	20.03	20	26.3	7.86	05/02/2018	18:00
27	CCV3	1	Water			22.6	2.00	05/02/2018	18:02
28	J2713-08	1	Solid	20.02	20	26.1	6.55	05/02/2018	18:04
29	CCV4	1	Water			22.8	12.00	05/02/2018	18:06

WORKLIST(Hardcopy Internal Chain)

LB95063

Worklist Name : corrosivity050218

Worklist ID : 111557

Date : 5/2/2018 3:30:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2018	Solid	J2670-02	Corrosivity	Cool 4 deg C	PSEG05	H63	TR-04-050218	05/02/2018	9045C
05/02/2018	Solid	J2679-02	Corrosivity	Cool 4 deg C	PSEG05	H63	EO-01-050218-A	05/02/2018	9045C
05/02/2018	Solid	J2679-04	Corrosivity	Cool 4 deg C	PSEG05	H63	EO-01-050218-B	05/02/2018	9045C
05/02/2018	Solid	J2679-06	Corrosivity	Cool 4 deg C	PSEG05	H63	EO-01-050218-C	05/02/2018	9045C
05/02/2018	Solid	J2681-02	Corrosivity	Cool 4 deg C	PSEG05	H62	NB-08-050218-A	05/02/2018	9045C
05/02/2018	Solid	J2681-04	Corrosivity	Cool 4 deg C	PSEG05	H62	NB-08-050218-B	05/02/2018	9045C
05/02/2018	Solid	J2686-02	Corrosivity	Cool 4 deg C	PSEG02	H62	GS-1-PAINT-CHIPS	05/02/2018	9045C
05/02/2018	Solid	J2686-04	Corrosivity	Cool 4 deg C	PSEG02	H62	GS-2-PAINT-CHIP-WIPES	05/02/2018	9045C
05/02/2018	Solid	J2695-02	Corrosivity	Cool 4 deg C	PSEG02	H62	CS-1-OILY-STONE	05/02/2018	9045C
05/02/2018	Solid	J2703-02	Corrosivity	Cool 4 deg C	PSEG01	H62	TP-1	05/02/2018	9045C
04/30/2018	Solid	J2706-04	Corrosivity	Cool 4 deg C	UNIO02	I52	001	04/30/2018	9045C
04/30/2018	Solid	J2706-05	Corrosivity	Cool 4 deg C	UNIO02	I52	002	04/30/2018	9045C
04/30/2018	Solid	J2706-06	Corrosivity	Cool 4 deg C	UNIO02	I52	003	04/30/2018	9045C
05/02/2018	Solid	J2710-03	Corrosivity	Cool 4 deg C	PSEG01	H52	1020	05/02/2018	9045C
05/02/2018	Solid	J2711-03	Corrosivity	Cool 4 deg C	PSEG01	H62	4215	05/02/2018	9045C
05/02/2018	Solid	J2712-03	Corrosivity	Cool 4 deg C	PSEG01	H62	0011	05/02/2018	9045C
05/02/2018	Solid	J2713-04	Corrosivity	Cool 4 deg C	PSEG01	H42	TP-1	05/02/2018	9045C
05/02/2018	Solid	J2713-08	Corrosivity	Cool 4 deg C	PSEG01	H42	TP-2	05/02/2018	9045C

Date/Time 05/02/18 8:35 PM

Received by: SP

Relinquished by: CP

Date/Time 05/02/18 6:15 PM

Received by: CP

Relinquished by: SP