



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95064

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2271         |
| buffer solution pH 7 yellow    | W2357         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2220         |
| buffer solution pH 7 yellow    | W2322         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID    | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|----------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1     | 1  | Water  |             |             | 22.5             | 4.01        | 05/02/2018 | 16:55     |
| 2   | CAL2     | 1  | Water  |             |             | 22.8             | 7.00        | 05/02/2018 | 16:57     |
| 3   | CAL3     | 1  | Water  |             |             | 22.6             | 10.03       | 05/02/2018 | 16:59     |
| 4   | ICV      | 1  | Water  |             |             | 22.8             | 7.00        | 05/02/2018 | 17:01     |
| 5   | CCV1     | 1  | Water  |             |             | 22.3             | 2.00        | 05/02/2018 | 17:03     |
| 6   | J2604-04 | 1  | Solid  | 20.02       | 20          | 21.4             | 7.12        | 05/02/2018 | 17:05     |
| 7   | CCV2     | 1  | Water  |             |             | 22.5             | 12.00       | 05/02/2018 | 17:07     |