



Analytical Summary Report

Analysis Method: 9041A

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95065

Slope : N/A

BalanceID: WC SC-4

pH Meter ID : N/A

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267
3.0-5.5 pH paper refill	W1934
6.0-8.0 pH paper refill	W2106
Test Paper,PH Short Range 9.0/10.0	W1939
6.0-8.0 pH paper refill	W2106
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			N/A	4.0	05/02/2018	17:20
2	CAL2	1	Water			N/A	7.0	05/02/2018	17:21
3	CAL3	1	Water			N/A	10.0	05/02/2018	17:22
4	ICV	1	Water			N/A	7.0	05/02/2018	17:23
5	CCV1	1	Water			N/A	2.0	05/02/2018	17:24
6	J2604-03	1	Solid	20.02	20	N/A	7.0	05/02/2018	17:25
7	CCV2	1	Water			N/A	12.0	05/02/2018	17:26