



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95067

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	05/02/2018	16:55
2	CAL2	1	Water			22.8	7.00	05/02/2018	16:57
3	CAL3	1	Water			22.6	10.03	05/02/2018	16:59
4	ICV	1	Water			22.8	7.00	05/02/2018	17:01
5	CCV1	1	Water			22.3	2.00	05/02/2018	17:03
6	J2696-01	1	Solid	20.02	20	24.4	5.12	05/02/2018	17:10
7	J2696-01DUP	1	Solid	20.03	20	24.5	5.13	05/02/2018	17:12
8	J2699-01	1	Solid	20.05	20	24.6	5.96	05/02/2018	17:14
9	CCV2	1	Water			22.5	12.00	05/02/2018	17:16

WORKLIST(Hardcopy Internal Chain)

LB 95067

Worklist Name : PH050218

Worklist ID : 111561

Date : 5/2/2018 4:38:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2018	Solid	J2696-01	pH	Cool 4 deg C	TULL02	I41	FK-PS-01.024	05/02/2018	9045C
05/02/2018	Solid	J2699-01	pH	Cool 4 deg C	TULL02	I43	FK-PS-01-024	05/02/2018	9045C

Date/Time

05/02/18 4:45 PM

Received by:

SP

Relinquished by:

SP

Date/Time

05/02/18 5:25 PM

Received by:

SP

Relinquished by:

SP