

ISM02.4


LB 950 77

INSTRUMENT ID: CV2

Linear

 μ Abs.:

40910



A= 0.0000e+000

B= 2.4644e-004

C= -3.6105e-002

Rho= 0.9999614

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.005	-0.005	126	0.000	126					10
0.05	0.050	0.060	0.010	392	0.0 %	392					-3
0.2	0.200	0.194	-0.006	934	0.0 %	934					1
2.5	2.500	2.531	0.031	10418	0.0 %	10418					0
5.0	5.000	4.987	-0.013	20384	0.0 %	20384					-1
7.5	7.500	7.436	-0.064	30322	0.0 %	30322					0
10.0	10.000	10.046	0.046	40910	0.0 %	40910					

AB
513118

LB95077
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	126	-		0 ISM02.4	PPB	5/3/2018 13:57	Std
	0.05 S0.05	392	-		0.05 ISM02.4	PPB	5/3/2018 13:59	Std
	0.2 S0.2	934	-		0.2 ISM02.4	PPB	5/3/2018 14:01	Std
	2.5 S2.5	10418	-		2.5 ISM02.4	PPB	5/3/2018 14:04	Std
	5 S5	20384	-		5 ISM02.4	PPB	5/3/2018 14:06	Std
	7.5 S7.5	30322	-		7.5 ISM02.4	PPB	5/3/2018 14:10	Std
	10 S10.0	40910	-		10 ISM02.4	PPB	5/3/2018 14:17	Std
ICV	ICV	15322	3.7398	-	ISM02.4	PPB	5/3/2018 14:20	SMPL
ICB	ICB	-36	-0.045	-	ISM02.4	PPB	5/3/2018 14:22	SMPL
CCV007	CCV007	20969	5.1314	-	ISM02.4	PPB	5/3/2018 14:25	SMPL
CCB007	CCB007	-60	-0.0509	-	ISM02.4	PPB	5/3/2018 14:27	SMPL
PB108849BL	PBW003	28	-0.0292	-	ISM02.4	PPB	5/3/2018 14:29	SMPL
J2516-01	MYAHG2	21	-0.0309	-	ISM02.4	PPB	5/3/2018 14:32	SMPL
J2516-02	MYAHG4	19	-0.0314	-	ISM02.4	PPB	5/3/2018 14:34	SMPL
J2516-03	MYAHH8	38	-0.0267	-	ISM02.4	PPB	5/3/2018 14:36	SMPL
J2516-04	MYAHH8D	6	-0.0346	-	ISM02.4	PPB	5/3/2018 14:38	SMPL
J2516-05	MYAHH8S	4359	1.0381	-	ISM02.4	PPB	5/3/2018 14:41	SMPL
J2516-06	MYAHJ0	7900	1.9107	-	ISM02.4	PPB	5/3/2018 14:43	SMPL
J2516-07	MYAHK0	231	0.0208	-	ISM02.4	PPB	5/3/2018 14:45	SMPL
J2516-08	MYAHM0	58011	14.2599	-	ISM02.4	PPB	5/3/2018 14:48	SMPL
J2516-09	MYAHN0	318	0.0423	-	ISM02.4	PPB	5/3/2018 14:52	SMPL
J2516-10	MYAHN2	474	0.0807	-	ISM02.4	PPB	5/3/2018 14:54	SMPL
PB108850BL	PBW003	343	0.0484	-	ISM02.4	PPB	5/3/2018 14:57	SMPL
J2517-01	MYAHG1	846	0.1724	-	ISM02.4	PPB	5/3/2018 14:59	SMPL
J2517-02	MYAHG3	635	0.1204	-	ISM02.4	PPB	5/3/2018 15:01	SMPL
J2517-03	MYAHG5	1258	0.2739	-	ISM02.4	PPB	5/3/2018 15:03	SMPL
J2517-04	MYAHG7	404	0.0635	-	ISM02.4	PPB	5/3/2018 15:06	SMPL
J2517-05	MYAHH3	30988	7.6005	-	ISM02.4	PPB	5/3/2018 15:08	SMPL
J2517-06	MYAHH5	367	0.0543	-	ISM02.4	PPB	5/3/2018 15:10	SMPL
J2517-07	MYAHH7	624	0.1177	-	ISM02.4	PPB	5/3/2018 15:13	SMPL
J2517-08	MYAHH7D	644	0.1226	-	ISM02.4	PPB	5/3/2018 15:15	SMPL
CCV008	CCV008	20679	5.06	-	ISM02.4	PPB	5/3/2018 15:17	SMPL
CCB008	CCB008	-56	-0.0499	-	ISM02.4	PPB	5/3/2018 15:19	SMPL
J2517-09	MYAHH7S	4474	1.0665	-	ISM02.4	PPB	5/3/2018 15:22	SMPL
J2517-10	MYAHH9	84347	20.7501	-	ISM02.4	PPB	5/3/2018 15:24	SMPL
J2517-11	MYAHJ1	39	-0.0265	-	ISM02.4	PPB	5/3/2018 15:29	SMPL
J2517-12	MYAHJ9	15621	3.8135	-	ISM02.4	PPB	5/3/2018 15:31	SMPL
J2517-13	MYAHK5	1268319	HIGH	-	ISM02.4	PPB	5/3/2018 15:33	
J2517-14	MYAHL1	37663	9.2454	-	ISM02.4	PPB	5/3/2018 15:35	SMPL
J2517-15	MYAHL9	469595	HIGH	-	ISM02.4	PPB	5/3/2018 15:40	
J2517-16	MYAHM1	964	0.2015	-	ISM02.4	PPB	5/3/2018 15:44	SMPL
J2517-17	MYAHM3	18	-0.0317	-	ISM02.4	PPB	5/3/2018 15:47	SMPL
J2517-18	MYAHM9	2101	0.4817	-	ISM02.4	PPB	5/3/2018 15:49	SMPL
J2517-19	MYAHN1	1605	0.3594	-	ISM02.4	PPB	5/3/2018 15:51	SMPL
J2517-20	MYAHM7	663	0.1273	-	ISM02.4	PPB	5/3/2018 15:54	SMPL
J2517-21	MYAHP1	52478	12.8964	-	ISM02.4	PPB	5/3/2018 15:56	SMPL

INSTRUMENT ID : CV1

CCV009	CCV009	20153	4.9303 -	ISM02.4	PPB	5/3/2018 15:58 SMPL
CCB009	CCB009	-113	-0.064 -	ISM02.4	PPB	5/3/2018 16:00 SMPL
J2516-08DLX2	MYAHM0	27809	6.8171 -	ISM02.4	PPB	5/3/2018 16:03 SMPL
J2517-10DLX5	MYAHH9	15687	3.8297 -	ISM02.4	PPB	5/3/2018 16:05 SMPL
J2517-13DLX50	MYAHK5	302974	74.6278 -	ISM02.4	PPB	5/3/2018 16:07 SMPL
J2517-15DLX10	MYAHL9	68214	16.7743 -	ISM02.4	PPB	5/3/2018 16:15 SMPL
J2517-21DLX2	MYAHP1	24984	6.1209 -	ISM02.4	PPB	5/3/2018 16:17 SMPL
J2517-13DL2X500	MYAHK5	32770	8.0396 -	ISM02.4	PPB	5/3/2018 16:19 SMPL
J2517-15DL2X50	MYAHL9	13017	3.1718 -	ISM02.4	PPB	5/3/2018 16:21 SMPL
CCV010	CCV010	20496	5.0149 -	ISM02.4	PPB	5/3/2018 16:24 SMPL
CCB010	CCB010	-191	-0.0832 -	ISM02.4	PPB	5/3/2018 16:26 SMPL