

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/3/2018 14:04

Test: ~~Total~~ ^{Reactive} CN ^{ug/d}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.913	0.0	0.099	
ICB1	-0.374	0.0	0.004	
CCV1	238.476	0.0	0.238	
CCB1	-0.878	0.0	0.004	
PB108877BL	-0.882	0.0	0.004	
J2706-04	-0.553	0.0	0.004	
J2706-04DUP	-0.556	0.0	0.004	
J2706-05	-0.920	0.0	0.004	
J2706-06	-0.730	0.0	0.004	
J2713-04	-0.906	0.0	0.004	
J2713-08	-1.035	0.0	0.004	
CCV2	255.373	0.0	0.254	
CCB2	-0.394	0.0	0.004	

N	13
Mean	44.887
SD	93.6709
CV%	208.68

Aquakem v. 7.2AQ1

Results from time period:

Thu May 03 13:53:14 2018

Thu May 03 13:56:38 2018

Sample Id	Sam/Ctr/c	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.7143	µg/l	5/3/2018 8:32:09	
5.OPPBCN	A	Reactive CN	P	3.8393	µg/l	5/3/2018 8:32:10	
10PPBCN	A	Reactive CN	P	9.6234	µg/l	5/3/2018 8:32:11	
50PPBCN	A	Reactive CN	P	50.3197	µg/l	5/3/2018 8:32:12	
100PPBCN	A	Reactive CN	P	100.9958	µg/l	5/3/2018 8:32:13	
250PPBCN	A	Reactive CN	P	252.2963	µg/l	5/3/2018 8:32:14	
500PPBCN	A	Reactive CN	P	498.6399	µg/l	5/3/2018 8:32:15	
ICV1	S	Reactive CN	P	96.9134	µg/l	5/3/2018 13:53:14	
ICB1	S	Reactive CN	P	-0.3742	µg/l	5/3/2018 13:53:15	
CCV1	S	Reactive CN	P	238.476	µg/l	5/3/2018 13:53:16	
CCB1	S	Reactive CN	P	-0.8777	µg/l	5/3/2018 13:53:17	
PB108877BL	S	Reactive CN	P	-0.8823	µg/l	5/3/2018 13:53:18	
J2706-04	S	Reactive CN	P	-0.5532	µg/l	5/3/2018 13:53:19	
J2706-04DUP	S	Reactive CN	P	-0.5565	µg/l	5/3/2018 13:53:20	
J2706-05	S	Reactive CN	P	-0.9205	µg/l	5/3/2018 13:53:21	
J2706-06	S	Reactive CN	P	-0.73	µg/l	5/3/2018 13:53:22	
J2713-04	S	Reactive CN	P	-0.9058	µg/l	5/3/2018 13:53:23	
J2713-08	S	Reactive CN	P	-1.035	µg/l	5/3/2018 13:53:24	
CCV2	S	Reactive CN	P	255.3726	µg/l	5/3/2018 13:56:37	
CCB2	S	Reactive CN	P	-0.3942	µg/l	5/3/2018 13:56:38	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

5/3/2018 14:17


Test Total CN

Accepted 5/3/2018 8:53

Factor 1022
Bias 0.005

Coeff. of det. 0.999950

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.7143	0.0000	
2	5.0PPBCN	0.008	3.8393	5.0000	
3	10PPBCN	0.014	9.6234	10.0000	
4	50PPBCN	0.054	50.3197	50.0000	
5	100PPBCN	0.103	100.9958	100.0000	
6	250PPBCN	0.251	252.2963	250.0000	
7	500PPBCN	0.492	498.6399	500.0000	