



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95087

Slope : 99.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			24.2	4.01	05/03/2018	15:05
2	CAL2	1	Water			24.5	7.00	05/03/2018	15:07
3	CAL3	1	Water			24.3	10.03	05/03/2018	15:09
4	ICV	1	Water			24.6	7.00	05/03/2018	15:11
5	CCV1	1	Water			24.3	2.00	05/03/2018	15:13
6	J2695-08	1	Water			24.5	7.70	05/03/2018	15:15
7	J2695-08DUP	1	Water			24.6	7.71	05/03/2018	15:17
8	J2729-02	1	Water			27.5	7.64	05/03/2018	15:19
9	CCV2	1	Water			24.5	12.00	05/03/2018	15:21

WORKLIST(Hardcopy Internal Chain)

LB 95087

Worklist Name : corrosivity050318

Worklist ID : 111620

Date : 5/3/2018 2:56:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/03/2018	Water	J2695-08	Corrosivity	Cool 4 deg C	PSEG02		CS-DRUM-2	05/03/2018	9040C
05/03/2018	Water	J2729-02	Corrosivity	Cool 4 deg C	PSEG01	J12	VB16119	05/03/2018	9040C

Date/Time 05/03/18 3:00 PM

Received by: SP

Relinquished by: CP

Date/Time 05/03/18 3:30 PM

Received by: CP

Relinquished by: SP