



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB95089

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: 0.0006 Slope: 0.7698 Regression: 0.999964

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB95089

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		1.79	0.000	0.392	0.392	0.508	05/03/2018	15:23
2	ICB		1	100	100		2.04	0.000	0.000	0.000	-0.001	05/03/2018	15:24
3	CCV1	0.5	1	100	100		2.31	0.000	0.392	0.392	0.508	05/03/2018	15:24
4	CCB1		1	100	100		2.26	0.000	0.000	0.000	-0.001	05/03/2018	15:25
5	PB108886BL		1	2.55	100	7.39	2.16	0.000	0.000	0.000	-0.001	05/03/2018	15:25
6	PB108886BS	20	1	2.53	100	7.11	1.60	0.000	0.360	0.360	0.467	05/03/2018	15:26
7	J2670-01		1	2.50	100	7.63	1.73	0.025	0.025	0.000	-0.001	05/03/2018	15:26
8	J2679-01		1	2.52	100	7.69	2.39	0.009	0.009	0.000	-0.001	05/03/2018	15:27
9	J2679-03		1	2.55	100	7.28	2.40	0.007	0.007	0.000	-0.001	05/03/2018	15:27
10	J2679-05		1	2.54	100	7.33	1.86	0.010	0.010	0.000	-0.001	05/03/2018	15:28
11	J2681-01		1	2.53	100	7.42	1.54	0.013	0.013	0.000	-0.001	05/03/2018	15:28
12	J2681-03		1	2.51	100	7.70	2.29	0.016	0.016	0.000	-0.001	05/03/2018	15:29
13	J2686-01		1	2.50	100	7.11	1.90	2.026	3.908	1.882	2.444	05/03/2018	15:29
14	J2686-03		1	2.52	100	7.15	1.87	0.181	0.181	0.000	-0.001	05/03/2018	15:30
15	CCV2	0.5	1	100	100		1.72	0.000	0.392	0.392	0.508	05/03/2018	15:30
16	CCB2		1	100	100		2.15	0.000	0.000	0.000	-0.001	05/03/2018	15:31
17	J2695-01		1	2.54	100	7.18	2.03	0.019	0.019	0.000	-0.001	05/03/2018	15:31
18	J2696-01		1	2.52	100	7.21	1.69	0.024	0.024	0.000	-0.001	05/03/2018	15:32
19	J2696-01DUP		1	2.51	100	7.06	1.82	0.023	0.023	0.000	-0.001	05/03/2018	15:32
20	J2696-01MSPre	40	2	2.50	100	7.08	2.20	0.026	0.291	0.265	0.343	05/03/2018	15:33
21	J2696-01MS2Ins	1284	40	2.51	100	7.10	1.77	0.021	0.615	0.594	0.771	05/03/2018	15:33
22	J2696-01MS3Post	40	2	2.52	100	7.26	1.59	0.029	0.310	0.281	0.364	05/03/2018	15:34
23	J2699-01		1	2.52	100	7.44	2.13	0.063	0.063	0.000	-0.001	05/03/2018	15:34
24	J2703-01		1	2.53	100	7.51	1.60	0.046	0.046	0.000	-0.001	05/03/2018	15:35
25	J2710-01		1	2.54	100	7.70	1.55	0.091	0.091	0.000	-0.001	05/03/2018	15:35
26	J2711-01		1	2.55	100	7.85	2.29	0.010	0.011	0.001	0.001	05/03/2018	15:36
27	CCV3	0.5	1	100	100		2.42	0.000	0.393	0.393	0.510	05/03/2018	15:36
28	CCB3		1	100	100		1.97	0.000	0.000	0.000	-0.001	05/03/2018	15:37
29	J2712-01		1	2.54	100	7.63	1.66	0.059	0.059	0.000	-0.001	05/03/2018	15:37
30	J2686-01		100	2.50	100	7.11	1.90	0.025	0.294	0.269	0.349	05/03/2018	15:38
31	CCV4	0.5	1	100	100		1.55	0.000	0.393	0.393	0.510	05/03/2018	15:38
32	CCB4		1	100	100		2.21	0.000	0.000	0.000	-0.001	05/03/2018	15:39