

LB 95092


7471B

INSTRUMENT ID: CVL

Linear

μ Abs.:

44229



A= 0.0000e+000

B= 2.2867e-004

C= -3.5565e-002

Rho= 0.9998173

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	43	0.000	43				
0.2	0.200	0.222	0.022	1127	0.0 %	1127				
2.5	2.500	2.535	0.035	11243	0.0 %	11243				
5.0	5.000	5.032	0.032	22161	0.0 %	22161				
7.5	7.500	7.358	-0.142	32331	0.0 %	32331				
10.0	10.000	10.078	0.078	44229	0.0 %	44229				

LB95092 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 May 2018 16:53:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	7471B	PPB	03 May 2018 16:58:50	S	Std
0.2 - 1	50.2	1127	-	0.2000	7471B	PPB	03 May 2018 17:01:07	S	Std
2.5 - 1	52.5	11243	-	2.5000	7471B	PPB	03 May 2018 17:03:25	S	Std
5.0 - 1	55	22161	-	5.0000	7471B	PPB	03 May 2018 17:05:42	S	Std
7.5 - 1	57.5	32331	-	7.5000	7471B	PPB	03 May 2018 17:08:00	S	Std
10.0 - 1	510	44229	-	10.0000	7471B	PPB	03 May 2018 17:10:17	S	Std
ICV28 - 1	ICV28	16252	3.6808	-	7471B	PPB	03 May 2018 17:13:08	U	SMPL
ICB28 - 1	ICB28	-198	-0.0808	-	7471B	PPB	03 May 2018 17:15:23	U	SMPL
CCV11 - 1	CCV11	22407	5.0883	-	7471B	PPB	03 May 2018 17:17:40	U	SMPL
CCB11 - 1	CCB11	-201	-0.0815	-	7471B	PPB	03 May 2018 17:19:56	U	SMPL
CRA - 1	CRA	1035	0.2011	-	7471B	PPB	03 May 2018 17:22:12	U	SMPL
HighStd - 1	HighStd	44243	10.0816	-	7471B	PPB	03 May 2018 17:24:27	U	SMPL
ChkStd - 1	ChkStd	29529	6.7169	-	7471B	PPB	03 May 2018 17:26:44	U	SMPL
PB108847BL - 1	PBS	-236	-0.0895	-	7471B	PPB	03 May 2018 17:29:02	U	SMPL
PB108847BS - 1	LCSS	15396	3.4851	-	7471B	PPB	03 May 2018 17:31:21	U	SMPL
J2669-01 - 1	SU-03-050118-A	5371	1.1926	-	7471B	PPB	03 May 2018 17:33:38	U	SMPL
J2669-03 - 1	SU-03-050118-B	722	0.1295	-	7471B	PPB	03 May 2018 17:35:56	U	SMPL
J2670-01 - 1	TR-04-050218	1035	0.2011	-	7471B	PPB	03 May 2018 17:38:13	U	SMPL
J2670-01DUP - 1	TR-04-050218DUP	966	0.1853	-	7471B	PPB	03 May 2018 17:40:31	U	SMPL
J2670-01MS - 1	TR-04-050218MS	18006	4.0819	-	7471B	PPB	03 May 2018 17:42:47	U	SMPL
CCV12 - 1	CCV12	22112	5.0209	-	7471B	PPB	03 May 2018 17:45:03	U	SMPL
CCB12 - 1	CCB12	-183	-0.0774	-	7471B	PPB	03 May 2018 17:47:19	U	SMPL
J2670-01MSD - 1	TR-04-050218MSD	18080	4.0989	-	7471B	PPB	03 May 2018 17:49:35	U	SMPL
J2679-01 - 1	EO-01-050218-A	914	0.1734	-	7471B	PPB	03 May 2018 17:51:51	U	SMPL
J2679-03 - 1	EO-01-050218-B	942	0.1798	-	7471B	PPB	03 May 2018 17:54:08	U	SMPL
J2679-05 - 1	EO-01-050218-C	1229	0.2455	-	7471B	PPB	03 May 2018 17:56:24	U	SMPL
J2681-01 - 1	NB-08-050218-A	1274	0.2558	-	7471B	PPB	03 May 2018 17:58:41	U	SMPL
J2681-03 - 1	NB-08-050218-B	1596	0.3294	-	7471B	PPB	03 May 2018 18:00:58	U	SMPL
J2686-01 - 1	GS-1-PAINT-CHIPS	42417	9.6641	-	7471B	PPB	03 May 2018 18:03:16	U	SMPL
J2686-03 - 1	GS-2-PAINT-CHIP-WIPES	972	0.1867	-	7471B	PPB	03 May 2018 18:05:33	U	SMPL
J2695-01 - 1	CS-1-OILY-STONE	4474	0.9875	-	7471B	PPB	03 May 2018 18:07:53	U	SMPL
CCV13 - 1	CCV13	22140	5.0273	-	7471B	PPB	03 May 2018 18:12:25	U	SMPL
CCB13 - 1	CCB13	-167	-0.0738	-	7471B	PPB	03 May 2018 18:14:41	U	SMPL
J2696-01 - 1	FK-PS-01.024	1241	0.2482	-	7471B	PPB	03 May 2018 18:16:59	U	SMPL
J2713-01 - 1	TP-1	1163	0.2304	-	7471B	PPB	03 May 2018 18:19:15	U	SMPL
J2713-05 - 1	TP-2	180	0.0056	-	7471B	PPB	03 May 2018 18:21:31	U	SMPL
PB108848BL - 1	PBS	-47	-0.0463	-	7471B	PPB	03 May 2018 18:23:47	U	SMPL
PB108848BS - 1	LCSS	15320	3.4677	-	7471B	PPB	03 May 2018 18:26:04	U	SMPL
J2699-01 - 1	FK-PS-01-024	458	0.0692	-	7471B	PPB	03 May 2018 18:28:21	U	SMPL
J2703-01 - 1	TP-1	15818	3.5816	-	7471B	PPB	03 May 2018 18:30:38	U	SMPL
J2705-01 - 1	EPE-F-GA(0-5)	938	0.1789	-	7471B	PPB	03 May 2018 18:32:55	U	SMPL
J2705-02 - 1	EPE-F-GA(5-10)	397	0.0552	-	7471B	PPB	03 May 2018 18:35:13	U	SMPL
J2705-03 - 1	EPE-MX-1(0-5)	7668	1.7179	-	7471B	PPB	03 May 2018 18:37:30	U	SMPL
CCV14 - 1	CCV14	22134	5.0259	-	7471B	PPB	03 May 2018 18:39:46	U	SMPL
CCB14 - 1	CCB14	-152	-0.0703	-	7471B	PPB	03 May 2018 18:42:02	U	SMPL
J2705-04 - 1	EPE-MX-1(5-10)	3197	0.6955	-	7471B	PPB	03 May 2018 18:44:19	U	SMPL
J2705-05 - 1	EPE-F-GA(35-37)	62	-0.0214	-	7471B	PPB	03 May 2018 18:46:35	U	SMPL
J2706-01 - 1	001	5623	1.2503	-	7471B	PPB	03 May 2018 18:48:52	U	SMPL
J2706-01DUP - 1	001DUP	5530	1.2290	-	7471B	PPB	03 May 2018 18:56:08	U	SMPL
J2706-01MS - 1	001MS	21194	4.8109	-	7471B	PPB	03 May 2018 18:58:25	U	SMPL
J2706-01MSD - 1	001MSD	20987	4.7636	-	7471B	PPB	03 May 2018 19:00:42	U	SMPL
J2706-02 - 1	002	5720	1.2724	-	7471B	PPB	03 May 2018 19:03:00	U	SMPL
J2706-03 - 1	003	1869	0.3918	-	7471B	PPB	03 May 2018 19:05:18	U	SMPL
J2709-01 - 1	WAYNE	383	0.0520	-	7471B	PPB	03 May 2018 19:07:35	U	SMPL
J2710-01 - 1	1020	12596	2.8448	-	7471B	PPB	03 May 2018 19:09:52	U	SMPL
CCV15 - 1	CCV15	21692	4.9248	-	7471B	PPB	03 May 2018 19:12:08	U	SMPL
CCB15 - 1	CCB15	-153	-0.0706	-	7471B	PPB	03 May 2018 19:14:24	U	SMPL
J2711-01 - 1	4215	3459	0.7554	-	7471B	PPB	03 May 2018 19:16:40	U	SMPL
J2712-01 - 1	0011	2077	0.4394	-	7471B	PPB	03 May 2018 19:18:56	U	SMPL
J2670-01LX5 - 1		118	-0.0086	-	7471B	PPB	03 May 2018 19:21:13	U	SMPL
J2670-01A - 1		18157	4.1165	-	7471B	PPB	03 May 2018 19:23:29	U	SMPL
J2706-01LX5 - 1		1395	0.2834	-	7471B	PPB	03 May 2018 19:30:30	U	SMPL
J2706-01A - 1		23537	5.3467	-	7471B	PPB	03 May 2018 19:32:46	U	SMPL
CCV16 - 1	CCV16	22188	5.0382	-	7471B	PPB	03 May 2018 19:35:04	U	SMPL
CCB16 - 1	CCB16	-186	-0.0781	-	7471B	PPB	03 May 2018 19:37:22	U	SMPL