

LB95116

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AS.

5/4/2018 12:13

Test: ^{Reactive} ~~Total~~ CN ^{Mg/d}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.150	0.0	0.098	
ICB1	0.430	0.0	0.004	
CCV1	244.900	0.0	0.244	
CCB1	-0.172	0.0	0.004	
PB108918BL	-0.911	0.0	0.003	
J2672-02	-0.729	0.0	0.003	
J2672-02DUP	-0.306	0.0	0.004	
J2722-03	0.123	0.0	0.004	
J2733-03	-0.083	0.0	0.004	
J2733-07	0.063	0.0	0.004	
J2733-11	-0.340	0.0	0.004	
PB108919BL	0.153	0.0	0.004	
J2695-08	0.046	0.0	0.004	
J2695-08DUP	-0.057	0.0	0.004	
CCV2	252.915	0.0	0.252	
CCB2	0.365	0.0	0.004	
J2729-02	0.327	0.0	0.004	
CCV3	252.858	0.0	0.252	
CCB3	0.060	0.0	0.004	

N	19
Mean	44.515
SD	94.1263
CV%	211.45

Aquakem v. 7.2AQ1

Results from time period:

Fri May 04 10:36:04 2018

Fri May 04 12:06:08 2018

AS

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.3465	µg/l	5/4/2018 10:36:04	
5.OPPBCN	A	Reactive CN	P	4.6086	µg/l	5/4/2018 10:36:05	
10PPBCN	A	Reactive CN	P	9.3075	µg/l	5/4/2018 10:36:06	
50PPBCN	A	Reactive CN	P	49.3193	µg/l	5/4/2018 10:36:07	
100PPBCN	A	Reactive CN	P	99.8865	µg/l	5/4/2018 10:36:08	
250PPBCN	A	Reactive CN	P	254.232	µg/l	5/4/2018 10:36:09	
500PPBCN	A	Reactive CN	P	497.9925	µg/l	5/4/2018 10:36:10	
ICV1	S	Reactive CN	P	96.1499	µg/l	5/4/2018 11:59:50	
ICB1	S	Reactive CN	P	0.4295	µg/l	5/4/2018 11:59:51	
CCV1	S	Reactive CN	P	244.9005	µg/l	5/4/2018 11:59:52	
CCB1	S	Reactive CN	P	-0.1719	µg/l	5/4/2018 11:59:53	
PB108918BL	S	Reactive CN	P	-0.9111	µg/l	5/4/2018 11:59:54	
J2672-02	S	Reactive CN	P	-0.7287	µg/l	5/4/2018 11:59:55	
J2672-02DUP	S	Reactive CN	P	-0.3056	µg/l	5/4/2018 11:59:56	
J2722-03	S	Reactive CN	P	0.1233	µg/l	5/4/2018 11:59:57	
J2733-03	S	Reactive CN	P	-0.0828	µg/l	5/4/2018 11:59:58	
J2733-07	S	Reactive CN	P	0.0628	µg/l	5/4/2018 11:59:59	
J2733-11	S	Reactive CN	P	-0.34	µg/l	5/4/2018 12:00:00	
PB108919BL	S	Reactive CN	P	0.1529	µg/l	5/4/2018 12:06:01	
J2695-08	S	Reactive CN	P	0.0457	µg/l	5/4/2018 12:06:02	
J2695-08DUP	S	Reactive CN	P	-0.0573	µg/l	5/4/2018 12:06:03	
CCV2	S	Reactive CN	P	252.9151	µg/l	5/4/2018 12:06:04	
CCB2	S	Reactive CN	P	0.3651	µg/l	5/4/2018 12:06:05	
J2729-02	S	Reactive CN	P	0.3274	µg/l	5/4/2018 12:06:06	
CCV3	S	Reactive CN	P	252.8583	µg/l	5/4/2018 12:06:07	
CCB3	S	Reactive CN	P	0.0596	µg/l	5/4/2018 12:06:08	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

5/4/2018 10:54


Test Total CN

Accepted 5/4/2018 10:54

Factor 1019
 Bias 0.004

Coeff. of det. 0.999887

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	0.0PPBCN	0.004	-0.3465	0.0000	
2	5.0PPBCN	0.008	4.6086	5.0000	
3	10PPBCN	0.013	9.3075	10.0000	
4	50PPBCN	0.052	49.3193	50.0000	
5	100PPBCN	0.102	99.8865	100.0000	
6	250PPBCN	0.253	254.2320	250.0000	
7	500PPBCN	0.492	497.9925	500.0000	