


7470A

LB 95122
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37932



A= 0.0000e+000

B= 2.6450e-004

C= -6.8196e-002

Rho= 0.9997461

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	290	0.000	290				
0.2	0.200	0.201	0.001	1017	0.0 %	1017				
2.5	2.500	2.551	0.051	9904	0.0 %	9904				
5.0	5.000	4.851	-0.149	18597	0.0 %	18597				
7.5	7.500	7.624	0.124	29080	0.0 %	29080				
10.0	10.000	9.965	-0.035	37932	0.0 %	37932				

LB95122 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 04 May 2018 13:51:10							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	290	-	0.0000	7470A	PPB	04 May 2018 13:53:59	S	Std
0.2 - 1	50.2	1017	-	0.2000	7470A	PPB	04 May 2018 13:56:07	S	Std
2.5 - 1	52.5	9904	-	2.5000	7470A	PPB	04 May 2018 13:58:15	S	Std
5.0 - 1	55	18597	-	5.0000	7470A	PPB	04 May 2018 14:00:23	S	Std
7.5 - 1	57.5	29080	-	7.5000	7470A	PPB	04 May 2018 14:02:31	S	Std
10.0 - 1	510	37932	-	10.0000	7470A	PPB	04 May 2018 14:04:40	S	Std
ICV30 - 1	ICV30	14880	3.8676	-	7470A	PPB	04 May 2018 14:07:21	U	SMPL
ICB30 - 1	ICB30	88	-0.0449	-	7470A	PPB	04 May 2018 14:09:27	U	SMPL
CCV26 - 1	CCV26	18528	4.8325	-	7470A	PPB	04 May 2018 14:11:34	U	SMPL
CCB26 - 1	CCB26	89	-0.0447	-	7470A	PPB	04 May 2018 14:13:41	U	SMPL
CRA - 1	CRA	1008	0.1984	-	7470A	PPB	04 May 2018 14:15:48	U	SMPL
HighStd - 1	HighStd	38017	9.9874	-	7470A	PPB	04 May 2018 14:17:55	U	SMPL
ChkStd - 1	ChkStd	25529	6.6843	-	7470A	PPB	04 May 2018 14:20:02	U	SMPL
PB108895BL - 1	PBW	211	-0.0124	-	7470A	PPB	04 May 2018 14:25:37	U	SMPL
PB108895BS - 1	LCSW	14250	3.7010	-	7470A	PPB	04 May 2018 14:27:43	U	SMPL
PB108895TB - 1	PB108839TB	59	-0.0526	-	7470A	PPB	04 May 2018 14:29:51	U	SMPL
J2670-02 - 1	TR-04-050218	108	-0.0396	-	7470A	PPB	04 May 2018 14:31:59	U	SMPL
J2679-02 - 1	EO-01-050218-A	162	-0.0253	-	7470A	PPB	04 May 2018 14:34:07	U	SMPL
J2679-04 - 1	EO-01-050218-B	128	-0.0343	-	7470A	PPB	04 May 2018 14:36:16	U	SMPL
J2679-06 - 1	EO-01-050218-C	138	-0.0317	-	7470A	PPB	04 May 2018 14:38:23	U	SMPL
CCV27 - 1	CCV27	18137	4.7291	-	7470A	PPB	04 May 2018 14:40:30	U	SMPL
CCB27 - 1	CCB27	117	-0.0372	-	7470A	PPB	04 May 2018 14:42:36	U	SMPL
J2681-02 - 1	NB-08-050218-A	155	-0.0272	-	7470A	PPB	04 May 2018 14:44:44	U	SMPL
J2681-04 - 1	NB-08-050218-B	163	-0.0251	-	7470A	PPB	04 May 2018 14:46:50	U	SMPL
J2686-02 - 1	GS-1-PAINT-CHIPS	163	-0.0251	-	7470A	PPB	04 May 2018 14:48:57	U	SMPL
J2686-04 - 1	GS-2-PAINT-CHIP-WIPES	174	-0.0222	-	7470A	PPB	04 May 2018 14:51:04	U	SMPL
J2695-02 - 1	CS-1-OILY-STONE	149	-0.0288	-	7470A	PPB	04 May 2018 14:53:11	U	SMPL
J2703-02 - 1	TP-1	42	-0.0571	-	7470A	PPB	04 May 2018 14:55:19	U	SMPL
J2706-04 - 1	001	66	-0.0507	-	7470A	PPB	04 May 2018 14:57:27	U	SMPL
J2706-05 - 1	002	65	-0.0510	-	7470A	PPB	04 May 2018 14:59:36	U	SMPL
J2706-06 - 1	003	77	-0.0478	-	7470A	PPB	04 May 2018 15:01:44	U	SMPL
J2710-03 - 1	1020	127	-0.0346	-	7470A	PPB	04 May 2018 15:03:51	U	SMPL
CCV28 - 1	CCV28	18345	4.7841	-	7470A	PPB	04 May 2018 15:05:57	U	SMPL
CCB28 - 1	CCB28	42	-0.0571	-	7470A	PPB	04 May 2018 15:08:04	U	SMPL
J2711-03 - 1	4215	83	-0.0462	-	7470A	PPB	04 May 2018 15:10:12	U	SMPL
J2712-03 - 1	0011	11	-0.0653	-	7470A	PPB	04 May 2018 15:12:19	U	SMPL
J2713-04 - 1	TP-1	55	-0.0536	-	7470A	PPB	04 May 2018 15:14:25	U	SMPL
J2713-08 - 1	TP-2	57	-0.0531	-	7470A	PPB	04 May 2018 15:16:33	U	SMPL
J2713-08DUP - 1	TP-2DUP	66	-0.0507	-	7470A	PPB	04 May 2018 15:18:40	U	SMPL
J2713-08MS - 1	TP-2MS	13849	3.5949	-	7470A	PPB	04 May 2018 15:20:48	U	SMPL
J2713-08MSD - 1	TP-2MSD	13474	3.4957	-	7470A	PPB	04 May 2018 15:22:56	U	SMPL
PB108896BL - 1	PBW	28	-0.0608	-	7470A	PPB	04 May 2018 15:29:39	U	SMPL
PB108896BS - 1	LCSW	14369	3.7324	-	7470A	PPB	04 May 2018 15:31:46	U	SMPL
J2695-07 - 1	CS-DRUM-2	8064	2.0648	-	7470A	PPB	04 May 2018 15:33:53	U	SMPL
CCV29 - 1	CCV29	18022	4.6987	-	7470A	PPB	04 May 2018 15:36:00	U	SMPL
CCB29 - 1	CCB29	135	-0.0325	-	7470A	PPB	04 May 2018 15:38:08	U	SMPL
J2695-07DUP - 1	CS-DRUM-2DUP	7148	1.8225	-	7470A	PPB	04 May 2018 15:40:16	U	SMPL
J2695-07MS - 1	CS-DRUM-2MS	23188	6.0651	-	7470A	PPB	04 May 2018 15:48:06	U	SMPL
J2695-07MSD - 1	CS-DRUM-2MSD	22314	5.8339	-	7470A	PPB	04 May 2018 15:52:43	U	SMPL
J2703-03 - 1	MANHOLE	1431	0.3103	-	7470A	PPB	04 May 2018 15:54:49	U	SMPL
J2729-01 - 1	VB16119	10749	2.7749	-	7470A	PPB	04 May 2018 15:56:57	U	SMPL
J2731-01 - 1	BP-LF1-SPW-1T-0518	315	0.0151	-	7470A	PPB	04 May 2018 15:59:05	U	SMPL
J2731-02 - 1	BP-LF1-SPW-1D-0518	189	-0.0182	-	7470A	PPB	04 May 2018 16:01:13	U	SMPL
J2731-03 - 1	BP-LF1-SPW-2T-0518	214	-0.0116	-	7470A	PPB	04 May 2018 16:03:21	U	SMPL
J2731-04 - 1	BP-LF1-SPW-2D-0518	227	-0.0082	-	7470A	PPB	04 May 2018 16:05:30	U	SMPL
J2731-05 - 1	BP-LF1-SPW-PB-0518	148	-0.0290	-	7470A	PPB	04 May 2018 16:07:37	U	SMPL
CCV30 - 1	CCV30	18886	4.9272	-	7470A	PPB	04 May 2018 16:09:44	U	SMPL
CCB30 - 1	CCB30	172	-0.0227	-	7470A	PPB	04 May 2018 16:11:51	U	SMPL
PB108897BL - 1	PBW	31	-0.0600	-	7470A	PPB	04 May 2018 16:13:59	U	SMPL
PB108897BS - 1	LCSW	14305	3.7155	-	7470A	PPB	04 May 2018 16:16:06	U	SMPL
PB108897TB - 1	PB108845TB	-53	-0.0822	-	7470A	PPB	04 May 2018 16:18:14	U	SMPL
J2703-04 - 1	MANHOLE	69	-0.0499	-	7470A	PPB	04 May 2018 16:20:21	U	SMPL
J2703-04DUP - 1	MANHOLEDUP	78	-0.0476	-	7470A	PPB	04 May 2018 16:22:28	U	SMPL
J2703-04MS - 1	MANHOLEMS	15426	4.0120	-	7470A	PPB	04 May 2018 16:24:36	U	SMPL
J2703-04MSD - 1	MANHOLEMSD	15256	3.9671	-	7470A	PPB	04 May 2018 16:26:44	U	SMPL
CCV31 - 1	CCV31	18618	4.8563	-	7470A	PPB	04 May 2018 16:28:52	U	SMPL
CCB31 - 1	CCB31	254	-0.0010	-	7470A	PPB	04 May 2018 16:32:36	U	SMPL
J2713-08LX5 - 1		116	-0.0375	-	7470A	PPB	04 May 2018 16:34:42	U	SMPL
J2713-08A - 1		13249	3.4362	-	7470A	PPB	04 May 2018 16:36:50	U	SMPL
J2695-07LX5 - 1		1563	0.3452	-	7470A	PPB	04 May 2018 16:38:57	U	SMPL
J2695-07A - 1		21072	5.5054	-	7470A	PPB	04 May 2018 16:41:04	U	SMPL
J2703-04LX5 - 1		-19	-0.0732	-	7470A	PPB	04 May 2018 16:43:12	U	SMPL
J2703-04A - 1		14593	3.7917	-	7470A	PPB	04 May 2018 16:45:21	U	SMPL
CCV32 - 1	CCV32	18279	4.7667	-	7470A	PPB	04 May 2018 16:47:28	U	SMPL
CCB32 - 1	CCB32	145	-0.0298	-	7470A	PPB	04 May 2018 16:49:36	U	SMPL