

ISM02.4


LB95149

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46936



A= 0.0000e+000

B= 2.1295e-004

C= -4.4706e-002

Rho= 0.9998678

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.004	-0.004	190	0.000	190					-
0.2	0.200	0.203	0.003	1165	0.0 %	1165					2
2.5	2.500	2.521	0.021	12049	0.0 %	12049					1
5.0	5.000	4.919	-0.081	23312	0.0 %	23312					-2
7.5	7.500	7.610	0.110	35948	0.0 %	35948					1
10.0	10.000	9.950	-0.050	46936	0.0 %	46936					-1

0.03
514118

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	190	-		0 ISM02.4	PPB	5/4/2018 20:16	Std
	0.2 S0.2	1165	-		0.2 ISM02.4	PPB	5/4/2018 20:18	Std
	2.5 S2.5	12049	-		2.5 ISM02.4	PPB	5/4/2018 20:20	Std
	5 S5	23312	-		5 ISM02.4	PPB	5/4/2018 20:22	Std
	7.5 S7.5	35948	-		7.5 ISM02.4	PPB	5/4/2018 20:25	Std
	10 S10.0	46936	-		10 ISM02.4	PPB	5/4/2018 20:33	Std
ICV	ICV	18641	3.9248 -		ISM02.4	PPB	5/4/2018 20:36	SMPL
ICB	ICB	30	-0.0383 -		ISM02.4	PPB	5/4/2018 20:38	SMPL
CCV033	CCV033	22913	4.8345 -		ISM02.4	PPB	5/4/2018 20:40	SMPL
CCB033	CCB033	21	-0.0402 -		ISM02.4	PPB	5/4/2018 20:43	SMPL
PB108949BL	PBS004	15	-0.0415 -		ISM02.4	PPB	5/4/2018 20:45	SMPL
J2606-19	MX2C81	18183	3.8273 -		ISM02.4	PPB	5/4/2018 20:47	SMPL
J2606-20	MX2C81D	17833	3.7527 -		ISM02.4	PPB	5/4/2018 20:54	SMPL
J2606-21	MX2C81S	29026	6.1362 -		ISM02.4	PPB	5/4/2018 20:59	SMPL
CCV034	CCV034	22365	4.7178 -		ISM02.4	PPB	5/4/2018 21:01	SMPL
CCB034	CCB034	-119	-0.07 -		ISM02.4	PPB	5/4/2018 21:03	SMPL