

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 05/07/2018

TEMP1 IN: 103 05/07/2018 07:10 TEMP OUT: 104 05/07/2018 08:20

TEMP2 IN: 104 05/07/2018 09:00 TEMP OUT: 104 05/07/2018 10:10

TEMP3 IN: 105 05/07/2018 12:00 TEMP OUT: 105 05/07/2018 13:35

TEMP4 IN: 105 05/07/2018 14:15 TEMP OUT: 105 05/07/2018 15:55

Run Number: LB95152

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB95152BL	LB95152BL	1.4164	1.4164	100	1.4164	1.4164	1.4164	0.0000	0
2	LB95152BS	LB95152BS	1.3821	1.3821	100	1.4319	1.4319	1.4319	0.0498	498
3	J2751-01	050318-SIDI-E-D-B	1.3945	1.3945	200	1.4064	1.4064	1.4064	0.0119	59.5
4	J2751-01DUP	050318-SIDI-E-D-BDUP	1.3909	1.3909	200	1.4030	1.4030	1.4030	0.0121	60.5
5	J2751-02	050318-SIDI-E-D-M	1.3817	1.3817	200	1.3909	1.3909	1.3909	0.0092	46
6	J2751-03	050318-SIDI-E-D-S	1.3954	1.3954	300	1.4069	1.4069	1.4069	0.0115	38.3
7	J2751-04	050318-SIDI-E-U-B	1.4014	1.4014	300	1.4149	1.4149	1.4149	0.0135	45
8	J2751-05	050318-SIDI-E-U-M	1.3881	1.3881	200	1.3987	1.3987	1.3987	0.0106	53
9	J2751-06	050318-SIDI-E-U-S	1.3962	1.3962	300	1.4022	1.4022	1.4022	0.0060	20
10	J2751-07	050318-JETW-E-D-B	1.3975	1.3975	300	1.4133	1.4133	1.4133	0.0158	52.7
11	J2751-08	050318-JETW-E-D-M	1.3907	1.3907	300	1.4033	1.4033	1.4033	0.0126	42
12	J2751-09	050318-JETW-E-D-S	1.4163	1.4163	300	1.4283	1.4283	1.4283	0.0120	40
13	J2751-10	050318-JETW-E-U-B	1.3890	1.3890	300	1.4027	1.4027	1.4027	0.0137	45.7
14	J2751-11	050318-JETW-E-U-M	1.3953	1.3953	200	1.4045	1.4045	1.4045	0.0092	46
15	J2751-11DUP	050318-JETW-E-U-MDUP	1.3871	1.3871	200	1.3965	1.3965	1.3965	0.0094	47

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 05/07/2018

TEMP1 IN: 103 05/07/2018 07:10 TEMP OUT: 104 05/07/2018 08:20

TEMP2 IN: 104 05/07/2018 09:00 TEMP OUT: 104 05/07/2018 10:10

TEMP3 IN: 105 05/07/2018 12:00 TEMP OUT: 105 05/07/2018 13:35

TEMP4 IN: 105 05/07/2018 14:15 TEMP OUT: 105 05/07/2018 15:55

Run Number: LB95152

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J2751-12	050318-JETW-E-U-S	1.3922	1.3922	300	1.4014	1.4014	1.4014	0.0092	30.7
17	J2751-13	050318-DBRI-F-D-B	1.4046	1.4046	300	1.4220	1.4220	1.4220	0.0174	58
18	J2751-14	050318-DBRI-F-D-M	1.3922	1.3922	300	1.4053	1.4053	1.4053	0.0131	43.7
19	J2751-15	050318-DBRI-F-D-S	1.3882	1.3882	300	1.3989	1.3989	1.3989	0.0107	35.7
20	J2751-16	050318-DBRI-F-U-B	1.4003	1.4003	300	1.4111	1.4111	1.4111	0.0108	36
21	J2751-17	050318-DBRI-F-U-M	1.4087	1.4087	300	1.4201	1.4201	1.4201	0.0114	38
22	J2751-18	050318-DBRI-F-U-S	1.3910	1.3910	300	1.4002	1.4002	1.4002	0.0092	30.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS050718-2

WorkList ID : 111702

Date : 5/7/2018 7:21:19 AM

21295152

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/10/2018	Water	J2751-01 C	TSS	Cool 4 deg C	HDR103	J22	050318-SIDI-E-D-B	05/03/2018	SM2540 D
05/10/2018	Water	J2751-02 C	TSS	Cool 4 deg C	HDR103	J22	050318-SIDI-E-D-M	05/03/2018	SM2540 D
05/10/2018	Water	J2751-03 C	TSS	Cool 4 deg C	HDR103	J22	050318-SIDI-E-D-S	05/03/2018	SM2540 D
05/10/2018	Water	J2751-04 C	TSS	Cool 4 deg C	HDR103	J22	050318-SIDI-E-U-B	05/03/2018	SM2540 D
05/10/2018	Water	J2751-05 C	TSS	Cool 4 deg C	HDR103	J22	050318-SIDI-E-U-M	05/03/2018	SM2540 D
05/10/2018	Water	J2751-06 C	TSS	Cool 4 deg C	HDR103	J22	050318-SIDI-E-U-S	05/03/2018	SM2540 D
05/10/2018	Water	J2751-07 C	TSS	Cool 4 deg C	HDR103	J22	050318-JETW-E-D-B	05/03/2018	SM2540 D
05/10/2018	Water	J2751-08 C	TSS	Cool 4 deg C	HDR103	J22	050318-JETW-E-D-M	05/03/2018	SM2540 D
05/10/2018	Water	J2751-09 C	TSS	Cool 4 deg C	HDR103	J22	050318-JETW-E-D-S	05/03/2018	SM2540 D
05/10/2018	Water	J2751-10 C	TSS	Cool 4 deg C	HDR103	J22	050318-JETW-E-U-B	05/03/2018	SM2540 D
05/10/2018	Water	J2751-11 C	TSS	Cool 4 deg C	HDR103	J22	050318-JETW-E-U-M	05/03/2018	SM2540 D
05/10/2018	Water	J2751-12 C	TSS	Cool 4 deg C	HDR103	J22	050318-JETW-E-U-S	05/03/2018	SM2540 D
05/10/2018	Water	J2751-13 C	TSS	Cool 4 deg C	HDR103	J22	050318-DBRI-F-D-B	05/03/2018	SM2540 D
05/10/2018	Water	J2751-14 C	TSS	Cool 4 deg C	HDR103	J22	050318-DBRI-F-D-M	05/03/2018	SM2540 D
05/10/2018	Water	J2751-15 C	TSS	Cool 4 deg C	HDR103	J22	050318-DBRI-F-D-S	05/03/2018	SM2540 D
05/10/2018	Water	J2751-16 C	TSS	Cool 4 deg C	HDR103	J22	050318-DBRI-F-U-B	05/03/2018	SM2540 D
05/10/2018	Water	J2751-17 C	TSS	Cool 4 deg C	HDR103	J22	050318-DBRI-F-U-M	05/03/2018	SM2540 D
05/10/2018	Water	J2751-18 C	TSS	Cool 4 deg C	HDR103	J22	050318-DBRI-F-U-S	05/03/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

05/07/18 09:00



Date/Time

Received by:

Relinquished by:

05/07/18 14:00

