



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95156

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	05/04/2018	16:35
2	CAL2	1	Water			22.3	7.00	05/04/2018	16:37
3	CAL3	1	Water			22.5	10.03	05/04/2018	16:39
4	ICV	1	Water			22.5	7.00	05/04/2018	16:41
5	CCV1	1	Water			22.4	2.00	05/04/2018	16:43
6	J2759-01	1	Water			21.6	8.24	05/04/2018	17:30
7	J2759-01DUP	1	Water			21.7	8.25	05/04/2018	17:32
8	J2764-01	1	Water			22.4	7.45	05/04/2018	17:34
9	CCV2	1	Water			22.6	12.00	05/04/2018	17:36

WORKLIST(Hardcopy Internal Chain)

LB 95156

WorkList Name : PH050418

WorkList ID : 111674

Date : 5/4/2018 3:58:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/04/2018	Water	J2759-01	pH	Cool 4 deg C	PSEG01	122	TP-2	05/04/2018	9040C
05/04/2018	Water	J2764-01	pH	Cool 4 deg C	PSEG01	123	WATER-WELL	05/04/2018	9040C

Date/Time

05/04/18 4:15 PM

Received by:

SP

Relinquished by:

CP

Date/Time

05/04/18 7:40 PM

Received by:

CP

Relinquished by:

SP