


7470A

LB95159
INSTRUMENT ID: CY1

Linear

 μ Abs.:

36506



A= 0.0000e+000

B= 2.7382e-004

C= -8.2654e-004

Rho= 0.9999351

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	54	0.000	54				
0.2	0.200	0.238	0.038	873	0.0 %	873				
2.5	2.500	2.452	-0.048	8956	0.0 %	8956				
5.0	5.000	4.943	-0.057	18055	0.0 %	18055				
7.5	7.500	7.558	0.058	27604	0.0 %	27604				
10.0	10.000	9.995	-0.005	36506	0.0 %	36506				

LB95159 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 May 2018 11:28:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7470A	PPB	07 May 2018 11:35:32	S	Std
0.2 - 1	50.2	873	-	0.2000	7470A	PPB	07 May 2018 11:37:40	S	Std
2.5 - 1	52.5	8956	-	2.5000	7470A	PPB	07 May 2018 11:39:49	S	Std
5.0 - 1	55	18055	-	5.0000	7470A	PPB	07 May 2018 11:41:57	S	Std
7.5 - 1	57.5	27604	-	7.5000	7470A	PPB	07 May 2018 11:44:05	S	Std
10.0 - 1	510	36506	-	10.0000	7470A	PPB	07 May 2018 11:46:12	S	Std
ICV31 - 1	ICV31	14017	3.8374	-	7470A	PPB	07 May 2018 11:48:47	U	SMPL
ICB31 - 1	ICB31	-126	-0.0353	-	7470A	PPB	07 May 2018 11:50:53	U	SMPL
CCV35 - 1	CCV35	18496	5.0638	-	7470A	PPB	07 May 2018 11:53:00	U	SMPL
CCB35 - 1	CCB35	-135	-0.0378	-	7470A	PPB	07 May 2018 11:55:06	U	SMPL
CRA - 1	CRA	711	0.1939	-	7470A	PPB	07 May 2018 11:57:14	U	SMPL
HighStd - 1	HighStd	36543	10.0055	-	7470A	PPB	07 May 2018 11:59:20	U	SMPL
ChkStd - 1	ChkStd	25693	7.0345	-	7470A	PPB	07 May 2018 12:01:27	U	SMPL
PB108940BL - 1	PBW	-241	-0.0668	-	7470A	PPB	07 May 2018 12:03:36	U	SMPL
PB108940BS - 1	LCSW	13382	3.6635	-	7470A	PPB	07 May 2018 12:05:45	U	SMPL
PB108940TB - 1	PB108890TB	-146	-0.0408	-	7470A	PPB	07 May 2018 12:07:52	U	SMPL
J2672-02 - 1	BU-02-050318	-36	-0.0107	-	7470A	PPB	07 May 2018 12:10:01	U	SMPL
J2722-03 - 1	OILY-STONES	-26	-0.0079	-	7470A	PPB	07 May 2018 12:12:09	U	SMPL
J2733-03 - 1	TP-1A-COMP	-22	-0.0069	-	7470A	PPB	07 May 2018 12:14:18	U	SMPL
J2733-07 - 1	TP-3AB-COMP	-48	-0.0140	-	7470A	PPB	07 May 2018 12:16:25	U	SMPL
CCV36 - 1	CCV36	18662	5.1093	-	7470A	PPB	07 May 2018 12:18:32	U	SMPL
CCB36 - 1	CCB36	-130	-0.0364	-	7470A	PPB	07 May 2018 12:20:38	U	SMPL
J2733-11 - 1	TP-5AB-COMP	-25	-0.0077	-	7470A	PPB	07 May 2018 12:22:46	U	SMPL
J2733-11DUP - 1	TP-5AB-COMPDUP	-14	-0.0047	-	7470A	PPB	07 May 2018 12:24:53	U	SMPL
J2733-11MS - 1	TP-5AB-COMPMS	13887	3.8018	-	7470A	PPB	07 May 2018 12:26:59	U	SMPL
J2733-11MSD - 1	TP-5AB-COMPMSD	13546	3.7084	-	7470A	PPB	07 May 2018 12:29:06	U	SMPL
PB108941BL - 1	PBW	-117	-0.0329	-	7470A	PPB	07 May 2018 12:31:13	U	SMPL
PB108941BS - 1	LCSW	13010	3.5616	-	7470A	PPB	07 May 2018 12:33:21	U	SMPL
PB108941TB - 1	PB108892TB	-112	-0.0315	-	7470A	PPB	07 May 2018 12:35:29	U	SMPL
J2695-08 - 1	CS-DRUM-2	-2	-0.0014	-	7470A	PPB	07 May 2018 12:37:37	U	SMPL
J2695-08DUP - 1	CS-DRUM-2DUP	27	0.0066	-	7470A	PPB	07 May 2018 12:39:46	U	SMPL
J2695-08MS - 1	CS-DRUM-2MS	13653	3.7377	-	7470A	PPB	07 May 2018 12:41:52	U	SMPL
CCV37 - 1	CCV37	18353	5.0247	-	7470A	PPB	07 May 2018 12:43:59	U	SMPL
CCB37 - 1	CCB37	-225	-0.0624	-	7470A	PPB	07 May 2018 12:46:06	U	SMPL
J2695-08MSD - 1	CS-DRUM-2MSD	13591	3.7207	-	7470A	PPB	07 May 2018 12:48:14	U	SMPL
J2729-02 - 1	VB16119	-98	-0.0277	-	7470A	PPB	07 May 2018 12:50:20	U	SMPL
J2733-11LX5 - 1		13	0.0027	-	7470A	PPB	07 May 2018 12:52:27	U	SMPL
J2733-11A - 1		13866	3.7960	-	7470A	PPB	07 May 2018 12:54:34	U	SMPL
J2695-08LX5 - 1		-117	-0.0329	-	7470A	PPB	07 May 2018 12:56:41	U	SMPL
J2695-08A - 1		13529	3.7037	-	7470A	PPB	07 May 2018 12:58:49	U	SMPL
CCV38 - 1	CCV38	18637	5.1024	-	7470A	PPB	07 May 2018 13:00:57	U	SMPL
CCB38 - 1	CCB38	-124	-0.0348	-	7470A	PPB	07 May 2018 13:03:06	U	SMPL