

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB95160BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP64262
calibration std. phosphate 0.5 ppm	WP64261
calibration std. phosphate 0.3 ppm	WP64260
calibration std. 0 ppm	WP64257
phosphate ICV std.	WP64266
phosphate CCV std.	WP64265
calibration std. phosphate 0.1 ppm	WP64259
calibration std. phosphate 0.05 ppm	WP64258
Combined reagent	WP64264
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0003Slope: 0.6486Regression: 0.999966

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/07/2018	11:04
2	CAL2	0.05	1	50	50	0.033	05/07/2018	11:05
3	CAL3	0.10	1	50	50	0.065	05/07/2018	11:05
4	CAL4	0.30	1	50	50	0.195	05/07/2018	11:06
5	CAL5	0.50	1	50	50	0.320	05/07/2018	11:06
6	CAL6	1.00	1	50	50	0.650	05/07/2018	11:07



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB95160

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.318	0.491	05/07/2018	11:07
2	ICB		1	50	50	0.000	0.000	05/07/2018	11:08
3	CCV1	0.50	1	50	50	0.321	0.495	05/07/2018	11:08
4	CCB1		1	50	50	0.000	0.000	05/07/2018	11:09
5	PB108978BL		1	1.01	50	0.000	0.000	05/07/2018	11:09
6	PB108978BS	0.50	1	1.00	50	0.314	0.485	05/07/2018	11:10
7	J2604-09		1	1.00	50	2.145	3.308	05/07/2018	11:11
8	J2604-09		50	1.00	50	0.411	0.634	05/07/2018	11:30
9	CCV2	0.50	1	50	50	0.320	0.494	05/07/2018	11:30
10	CCB2		1	50	50	0.000	0.000	05/07/2018	11:31