



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95162

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	05/07/2018	09:10
2	J2695-08		1	>100.0	>212.0	>212.0	05/07/2018	09:32
3	J2695-08DUP		1	>100.0	>212.0	>212.0	05/07/2018	09:50
4	J2729-02		1	>100.0	>212.0	>212.0	05/07/2018	10:15
5	J2759-02		1	>100.0	>212.0	>212.0	05/07/2018	10:40
6	J2764-02		1	>100.0	>212.0	>212.0	05/07/2018	11:02

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB 95162

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP2-050718

WorkList ID : 111718

Date : 5/7/2018 8:15:24 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/08/2018	Water	J2695-08 B	Flash Point	Cool 4 deg C	PSEG02		CS-DRUM-2	05/03/2018	1010A
05/08/2018	Water	J2729-02 C	Flash Point	Cool 4 deg C	PSEG01	J12	VB16119	05/03/2018	1010A
05/09/2018	Water	J2759-02 C	Flash Point	Cool 4 deg C	PSEG01	I22	TP-2	05/04/2018	1010A
05/09/2018	Water	J2764-02 C	Flash Point	Cool 4 deg C	PSEG01	I23	WATER-WELL	05/04/2018	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: