



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95163

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	82.4	05/07/2018	11:30
2	J2735-03		1	>100.0	>212.0	>212.0	05/07/2018	11:52
3	J2735-03DUP		1	>100.0	>212.0	>212.0	05/07/2018	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-050718

WorkList ID : 111716

Date : 5/7/2018 8:14:35 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/17/2018	Solid	J2735-03	E	Cool 4 deg C	ARBO01	J41	OIL-DRUM	05/03/2018	1010A

Date/Time 05/07/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 05/07/18 12:20 PM
 Received by: CP
 Relinquished by: KS