



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95183

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	05/07/2018	16:15
2	CAL2	1	Water			21.3	7.00	05/07/2018	16:17
3	CAL3	1	Water			21.5	10.03	05/07/2018	16:19
4	ICV	1	Water			21.6	7.00	05/07/2018	16:21
5	CCV1	1	Water			21.4	2.00	05/07/2018	16:23
6	J2787-01	1	Water			21.5	7.38	05/07/2018	16:25
7	J2787-01DUP	1	Water			21.6	7.39	05/07/2018	16:27
8	CCV2	1	Water			21.5	12.00	05/07/2018	16:29

WORKLIST(Hardcopy Internal Chain)

LB 95183

WorkList Name : PH050718

WorkList ID : 111751

Date : 5/7/2018 3:48:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2018	Water	J2787-01	pH	Cool 4 deg C	PSEG01	I31	V23608-VB102	05/07/2018	9040C

Date/Time 05/07/18 4:00 PM

Received by: SP

Relinquished by: CP

Date/Time 05/07/18 4:35 PM

Received by: CP

Relinquished by: SP