

===== LB 95194 =====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AS5/8/2018 14:45
-----Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.604	0.0	0.097	
ICB1	-0.537	0.0	0.003	
CCV1	242.702	0.0	0.236	
CCB1	-0.226	0.0	0.004	
PB109036BL	-0.298	0.0	0.004	
PB109036BS	98.163	0.0	0.098	
J2671-01	0.869	0.0	0.005	
J2671-01DUP	-0.055	0.0	0.004	
J2671-01MS	34.734	0.0	0.037	
J2671-01MSD	34.224	0.0	0.037	
J2671-03	-0.040	0.0	0.004	
J2674-01	0.988	0.0	0.005	
J2676-03	0.297	0.0	0.004	
J2676-05	0.490	0.0	0.004	
CCV2	243.464	0.0	0.237	
CCB2	-0.690	0.0	0.003	
J2676-07	1.108	0.0	0.005	
J2676-09	1.236	0.0	0.005	
J2759-04	-0.263	0.0	0.004	
J2759-07	0.049	0.0	0.004	
J2676-01	-0.202	0.0	0.004	
CCV3	258.834	0.0	0.252	
CCB3	-0.433	0.0	0.004	

N	23
Mean	44.001
SD	85.8620
CV%	195.14

AS.

Aquakem v. 7.2AQ1

Results from time period:

Tue May 08 14:15:38 2018

Tue May 08 14:44:57 2018

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.8799	µg/l	5/8/2018 9:51:23	
5.0PPBCN	A	Total CN	P	5.1561	µg/l	5/8/2018 9:51:24	
10PPBCN	A	Total CN	P	9.8503	µg/l	5/8/2018 9:51:25	
50PPBCN	A	Total CN	P	49.6034	µg/l	5/8/2018 9:51:26	
100PPBCN	A	Total CN	P	100.4855	µg/l	5/8/2018 9:51:27	
250PPBCN	A	Total CN	P	251.6812	µg/l	5/8/2018 9:51:28	
500PPBCN	A	Total CN	P	499.1034	µg/l	5/8/2018 9:51:29	
ICV1	S	Total CN	P	97.6039	µg/l	5/8/2018 14:15:38	
ICB1	S	Total CN	P	-0.5366	µg/l	5/8/2018 14:15:39	
CCV1	S	Total CN	P	242.7017	µg/l	5/8/2018 14:15:40	
CCB1	S	Total CN	P	-0.2259	µg/l	5/8/2018 14:15:41	
PB109036BL	S	Total CN	P	-0.2977	µg/l	5/8/2018 14:15:42	
PB109036BS	S	Total CN	P	98.1629	µg/l	5/8/2018 14:15:43	
J2671-01	S	Total CN	P	0.8685	µg/l	5/8/2018 14:15:44	
J2671-01DUP	S	Total CN	P	-0.0547	µg/l	5/8/2018 14:15:45	
J2671-01MS	S	Total CN	P	34.7342	µg/l	5/8/2018 14:15:46	
J2671-01MSD	S	Total CN	P	34.2242	µg/l	5/8/2018 14:15:47	
J2671-03	S	Total CN	P	-0.0403	µg/l	5/8/2018 14:15:48	
J2674-01	S	Total CN	P	0.9879	µg/l	5/8/2018 14:23:13	
J2676-03	S	Total CN	P	0.2973	µg/l	5/8/2018 14:23:14	
J2676-05	S	Total CN	P	0.4902	µg/l	5/8/2018 14:23:15	
CCV2	S	Total CN	P	243.4636	µg/l	5/8/2018 14:23:16	
CCB2	S	Total CN	P	-0.6895	µg/l	5/8/2018 14:23:17	
J2676-07	S	Total CN	P	1.1082	µg/l	5/8/2018 14:23:18	
J2676-09	S	Total CN	P	1.2362	µg/l	5/8/2018 14:23:19	
J2759-04	S	Total CN	P	-0.2631	µg/l	5/8/2018 14:23:20	
J2759-07	S	Total CN	P	0.0493	µg/l	5/8/2018 14:23:21	
J2676-01	S	Total CN	P	-0.2023	µg/l	5/8/2018 14:44:55	
CCV3	S	Total CN	P	258.8336	µg/l	5/8/2018 14:44:56	
CCB3	S	Total CN	P	-0.4332	µg/l	5/8/2018 14:44:57	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS.

5/8/2018 9:52


Test Total CN

Accepted 5/8/2018 9:52

Factor 1045
 Bias 0.004

Coeff. of det. 0.999976

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.8799	0.0000	
2	5.0PPBCN	0.009	5.1561	5.0000	
3	10PPBCN	0.013	9.8503	10.0000	
4	50PPBCN	0.051	49.6034	50.0000	
5	100PPBCN	0.100	100.4855	100.0000	
6	250PPBCN	0.245	251.6812	250.0000	
7	500PPBCN	0.482	499.1034	500.0000	