

7470A


LB95206

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35457



A= 0.0000e+000

B= 2.7985e-004

C= -1.1436e-002

Rho= 0.9991384

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-11	0.000	-11				
0.2	0.200	0.249	0.049	929	0.0 %	929				
2.5	2.500	2.562	0.062	9196	0.0 %	9196				
5.0	5.000	4.745	-0.255	16997	0.0 %	16997				
7.5	7.500	7.748	0.248	27727	0.0 %	27727				
10.0	10.000	9.911	-0.089	35457	0.0 %	35457				

LB95206 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 May 2018 08:16:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.0000	7470A	PPB	09 May 2018 08:19:33	S	Std
0.2 - 1	50.2	929	-	0.2000	7470A	PPB	09 May 2018 08:21:41	S	Std
2.5 - 1	52.5	9196	-	2.5000	7470A	PPB	09 May 2018 08:23:50	S	Std
5.0 - 1	55	16997	-	5.0000	7470A	PPB	09 May 2018 08:25:58	S	Std
7.5 - 1	57.5	27727	-	7.5000	7470A	PPB	09 May 2018 08:28:06	S	Std
10.0 - 1	510	35457	-	10.0000	7470A	PPB	09 May 2018 08:32:43	S	Std
ICV36 - 1	ICV36	14059	3.9229	-	7470A	PPB	09 May 2018 08:35:17	U	SMPL
ICB36 - 1	ICB36	-143	-0.0515	-	7470A	PPB	09 May 2018 08:37:23	U	SMPL
CCV65 - 1	CCV65	17494	4.8842	-	7470A	PPB	09 May 2018 08:39:31	U	SMPL
CCB65 - 1	CCB65	-142	-0.0512	-	7470A	PPB	09 May 2018 08:41:37	U	SMPL
CRA - 1	CRA	897	0.2396	-	7470A	PPB	09 May 2018 08:43:44	U	SMPL
HighStd - 1	HighStd	34966	9.7736	-	7470A	PPB	09 May 2018 08:45:51	U	SMPL
ChkStd - 1	ChkStd	25613	7.1562	-	7470A	PPB	09 May 2018 08:47:58	U	SMPL
PB109053BL - 1	PBW	-224	-0.0741	-	7470A	PPB	09 May 2018 08:50:06	U	SMPL
PB109053BS - 1	LCSW	12122	3.3808	-	7470A	PPB	09 May 2018 08:52:15	U	SMPL
PB109053TB - 1	PB109008TB	-105	-0.0408	-	7470A	PPB	09 May 2018 08:54:23	U	SMPL
J2671-02 - 1	AU-02-050718-A	-48	-0.0249	-	7470A	PPB	09 May 2018 08:56:31	U	SMPL
J2671-02DUP - 1	AU-02-050718-ADUP	-18	-0.0165	-	7470A	PPB	09 May 2018 08:58:39	U	SMPL
J2671-02MS - 1	AU-02-050718-AMS	12703	3.5434	-	7470A	PPB	09 May 2018 09:00:48	U	SMPL
J2671-02MSD - 1	AU-02-050718-AMSD	12849	3.5843	-	7470A	PPB	09 May 2018 09:02:54	U	SMPL
CCV66 - 1	CCV66	17417	4.8626	-	7470A	PPB	09 May 2018 09:05:00	U	SMPL
CCB66 - 1	CCB66	-122	-0.0456	-	7470A	PPB	09 May 2018 09:07:06	U	SMPL
J2671-04 - 1	AU-02-050718-B	-35	-0.0212	-	7470A	PPB	09 May 2018 09:09:14	U	SMPL
J2674-02 - 1	HD-02-050718	-8	-0.0137	-	7470A	PPB	09 May 2018 09:11:21	U	SMPL
J2676-02 - 1	CL-01-050718-A	-4	-0.0126	-	7470A	PPB	09 May 2018 09:13:27	U	SMPL
J2676-04 - 1	CL-01-050718-B	-24	-0.0182	-	7470A	PPB	09 May 2018 09:15:35	U	SMPL
J2676-06 - 1	CL-01-050718-C	-23	-0.0179	-	7470A	PPB	09 May 2018 09:17:42	U	SMPL
J2676-08 - 1	CL-01-050718-D	-6	-0.0131	-	7470A	PPB	09 May 2018 09:19:50	U	SMPL
J2676-10 - 1	CL-01-050718-E	-15	-0.0156	-	7470A	PPB	09 May 2018 09:21:58	U	SMPL
J2759-06 - 1	TP-2A-COMP	-14	-0.0154	-	7470A	PPB	09 May 2018 09:24:06	U	SMPL
J2759-09 - 1	TP-2B-COMP	218	0.0496	-	7470A	PPB	09 May 2018 09:26:15	U	SMPL
J2780-06 - 1	SOIL	-18	-0.0165	-	7470A	PPB	09 May 2018 09:28:21	U	SMPL
CCV67 - 1	CCV67	17197	4.8011	-	7470A	PPB	09 May 2018 09:30:28	U	SMPL
CCB67 - 1	CCB67	-134	-0.0489	-	7470A	PPB	09 May 2018 09:32:35	U	SMPL
J2783-03 - 1	BS-CONCRETE-RO-277	-16	-0.0159	-	7470A	PPB	09 May 2018 09:34:43	U	SMPL
J2783-06 - 1	BS-OILY-STONE-RO-208	-21	-0.0173	-	7470A	PPB	09 May 2018 09:36:50	U	SMPL
J2784-02 - 1	RO-160	-4	-0.0126	-	7470A	PPB	09 May 2018 09:38:57	U	SMPL
PB109054BL - 1	PBW	-5	-0.0128	-	7470A	PPB	09 May 2018 09:41:04	U	SMPL
PB109054BS - 1	LCSW	14211	3.9654	-	7470A	PPB	09 May 2018 09:43:12	U	SMPL
PB109054TB - 1	PB109041TB	-124	-0.0461	-	7470A	PPB	09 May 2018 09:45:19	U	SMPL
J2787-02 - 1	V23608-VB102	-4	-0.0126	-	7470A	PPB	09 May 2018 09:47:27	U	SMPL
J2787-02DUP - 1	V23608-VB102DUP	-21	-0.0173	-	7470A	PPB	09 May 2018 09:49:35	U	SMPL
J2787-02MS - 1	V23608-VB102MS	12975	3.6196	-	7470A	PPB	09 May 2018 09:51:43	U	SMPL
J2787-02MSD - 1	V23608-VB102MSD	12974	3.6193	-	7470A	PPB	09 May 2018 09:53:50	U	SMPL
CCV68 - 1	CCV68	17274	4.8226	-	7470A	PPB	09 May 2018 09:55:57	U	SMPL
CCB68 - 1	CCB68	-149	-0.0531	-	7470A	PPB	09 May 2018 09:58:04	U	SMPL
PB109055BL - 1	PBW	-35	-0.0212	-	7470A	PPB	09 May 2018 10:00:12	U	SMPL
PB109055BS - 1	LCSW	11995	3.3453	-	7470A	PPB	09 May 2018 10:02:19	U	SMPL
J2787-01 - 1	V23608-VB102	13613	3.7981	-	7470A	PPB	09 May 2018 10:04:26	U	SMPL
J2787-01DUP - 1	V23608-VB102DUP	11859	3.3072	-	7470A	PPB	09 May 2018 10:06:33	U	SMPL
J2787-01MS - 1	V23608-VB102MS	25322	7.0748	-	7470A	PPB	09 May 2018 10:09:11	U	SMPL
J2787-01MSD - 1	V23608-VB102MSD	24801	6.9290	-	7470A	PPB	09 May 2018 10:11:19	U	SMPL
J2677-04 - 1	OK-01-050718	-180	-0.0618	-	7470A	PPB	09 May 2018 10:13:28	U	SMPL
CCV69 - 1	CCV69	16657	4.6499	-	7470A	PPB	09 May 2018 10:15:38	U	SMPL
CCB69 - 1	CCB69	-90	-0.0366	-	7470A	PPB	09 May 2018 10:17:47	U	SMPL
J2671-02LX5 - 1		-10	-0.0142	-	7470A	PPB	09 May 2018 10:19:54	U	SMPL
J2671-02A - 1		13164	3.6724	-	7470A	PPB	09 May 2018 10:22:02	U	SMPL
J2787-02LX5 - 1		-101	-0.0397	-	7470A	PPB	09 May 2018 10:24:09	U	SMPL
J2787-02A - 1		13339	3.7214	-	7470A	PPB	09 May 2018 10:26:16	U	SMPL
J2787-01LX5 - 1		2845	0.7847	-	7470A	PPB	09 May 2018 10:28:23	U	SMPL
J2787-01A - 1		26372	7.3686	-	7470A	PPB	09 May 2018 10:30:30	U	SMPL
CCV70 - 1	CCV70	17195	4.8005	-	7470A	PPB	09 May 2018 10:32:37	U	SMPL
CCB70 - 1	CCB70	-124	-0.0461	-	7470A	PPB	09 May 2018 10:34:46	U	SMPL