



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95209

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/08/2018	17:50
2	CAL2	1	Water			21.3	7.00	05/08/2018	17:52
3	CAL3	1	Water			21.4	10.03	05/08/2018	17:54
4	ICV	1	Water			21.5	7.00	05/08/2018	17:56
5	CCV1	1	Water			21.3	2.00	05/08/2018	18:34
6	J2803-07	1	Water			21.3	6.72	05/08/2018	18:36
7	J2803-07DUP	1	Water			21.4	6.73	05/08/2018	18:38
8	J2811-14	1	Water			18.0	7.02	05/08/2018	18:40
9	CCV2	1	Water			21.6	12.00	05/08/2018	18:42

WORKLIST(Hardcopy Internal Chain)

LB 95209

WorkList Name : corrosivity J2803

WorkList ID : 111801

Date : 5/8/2018 4:50:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/08/2018	Water	J2803-07	Corrosivity	Cool 4 deg C	PSEG02	I22	OILY-WATER	05/08/2018	9040C
05/08/2018	Water	J2811-14	Corrosivity	Cool 4 deg C	PSEG01	I33	TP-6	05/08/2018	9040C

Date/Time 05/08/18 5:45 PM

Received by: SP

Relinquished by: CP

Date/Time 05/08/18 6:50 PM

Received by: CP

Relinquished by: SP